



284 Sheffield Street, Mountainside, NJ 07092 Phone: 908 789 8900 Fax: 908 789 8922

8C

SEMIVOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH

Lab Code: CHEM Case No.: 05252 SAS No.: 05252 SDG NO.: 05252

EPA Sample No.: SSTDCCC040 Date Analyzed: 11/06/2023

Lab File ID: BF136149.D Time Analyzed: 12:01

Instrument ID: BNA_F GC Column: DB-UI ID: 0.18 (mm)

	IS4 (PHN) AREA #	RT #	IS5 (CRY) AREA #	RT #	IS6 (PRY) AREA #	RT #
12 HOUR STD	390010	11.351	198566	14.01	195040	15.492
UPPER LIMIT	780020	11.851	397132	14.51	390080	15.992
LOWER LIMIT	195005	10.851	99283	13.51	97520	14.992
EPA SAMPLE NO.						
01 WC-2MS	266358	11.35	168860	14.01	162522	15.50
02 WC-2MSD	270586	11.35	171083	14.01	163710	15.50

IS4 (PHN) = Phenanthrene-d10

IS5 (CRY) = Chrysene-d12

IS6 (PRY) = Perylene-d12

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT UPPER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.