



284 Sheffield Street, Mountainside, NJ 07092 Phone: 908 789 8900 Fax: 908 789 8922

8C

SEMIVOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH

Lab Code: CHEM Case No.: 05252 SAS No.: 05252 SDG NO.: 05252

EPA Sample No.: SSTDCCC040 Date Analyzed: 11/10/2023

Lab File ID: BM042685.D Time Analyzed: 11:22

Instrument ID: BNA_M GC Column: ZB-GR ID: 0.25 (mm)

01

	IS4 (PHN) AREA #	RT #	IS5 (CRY) AREA #	RT #	IS6 (PRY) AREA #	RT #
12 HOUR STD	472580	17.257	366899	21.445	365656	23.821
UPPER LIMIT	945160	17.757	733798	21.945	731312	24.321
LOWER LIMIT	236290	16.757	183450	20.945	182828	23.321
EPA SAMPLE NO.						
WASTE	307838	17.26	232800	21.44	261247	23.82

IS4 (PHN) = Phenanthrene-d10

IS5 (CRY) = Chrysene-d12

IS6 (PRY) = Perylene-d12

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT UPPER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.