



284 Sheffield Street, Mountainside, NJ 07092 Phone: 908 789 8900 Fax: 908 789 8922

Instrument ID: MSVOA_Y

Daily Analysis Runlog For Sequence/QC Batch ID # VY103123

Review By	John Carlone	Review On	11/1/2023 10:05:25 AM
Supervise By	Mahesh Dadoda	Supervise On	11/1/2023 4:45:21 PM
SubDirectory	VY103123	HP Acquire Method	HP Processing Method 82y103123s.m
STD. NAME	STD REF.#		
Tune/Reschk	VP123921		
Initial Calibration Stds	VP123922,VP123923,VP123924,VP123925,VP123926,VP123927		
CCC	VP123940,VP123941,LOD-VP123954,VP123955		
Internal Standard/PEM	VP123883		
ICV/I.BLK	VP123930		
Surrogate Standard			
MS/MSD Standard			
LCS Standard			

Sr#	SampleId	Data File Name	Date-Time	Operator	Status
1	BFB	VY016140.D	31 Oct 2023 08:42	SY/MD	Ok
2	VSTDCCC050	VY016141.D	31 Oct 2023 09:54	SY/MD	Not Ok
3	VSTDICC005	VY016142.D	31 Oct 2023 12:26	SY/MD	Ok,M
4	VSTDICC010	VY016143.D	31 Oct 2023 12:49	SY/MD	Ok,M
5	VSTDICC020	VY016144.D	31 Oct 2023 13:12	SY/MD	Ok,M
6	VSTDICCC050	VY016145.D	31 Oct 2023 13:37	SY/MD	Ok,M
7	VSTDICC100	VY016146.D	31 Oct 2023 14:13	SY/MD	Ok,M
8	VSTDICC150	VY016147.D	31 Oct 2023 14:40	SY/MD	Ok,M
9	VSTDICV050	VY016148.D	31 Oct 2023 15:03	SY/MD	Ok,M
10	VY1031SBL01	VY016149.D	31 Oct 2023 16:28	SY/MD	Ok
11	VY1031SBS01	VY016150.D	31 Oct 2023 16:51	SY/MD	Ok,M
12	VY1031SBSD01	VY016151.D	31 Oct 2023 17:13	SY/MD	Ok,M
13	O4699-01 2.5PPB	VY016152.D	31 Oct 2023 17:36	SY/MD	Ok,M
14	O4699-01 4.0PPB	VY016153.D	31 Oct 2023 17:59	SY/MD	Ok,M
15	O5136-01	VY016154.D	31 Oct 2023 18:23	SY/MD	Ok
16	O5107-23	VY016155.D	31 Oct 2023 18:47	SY/MD	Ok
17	VSTDCCC050	VY016156.D	31 Oct 2023 19:32	SY/MD	Ok,M

M : Manual Integration



284 Sheffield Street, Mountainside, NJ 07092 Phone: 908 789 8900 Fax: 908 789 8922

Instrument ID: MSVOA_Y

Daily Analysis Runlog For Sequence/QC Batch ID # VY110723

Review By	John Carlone	Review On	11/8/2023 11:48:05 AM		
Supervise By	Mahesh Dadoda	Supervise On	11/8/2023 3:22:06 PM		
SubDirectory	VY110723	HP Acquire Method	MSVOA_Y	HP Processing Method	82y103123s.m
STD. NAME		STD REF.#			
Tune/Reschk Initial Calibration Stds		VP124090			
CCC		VP124091,VP124092			
Internal Standard/PEM		VP123883			
ICV/I.BLK					
Surrogate Standard					
MS/MSD Standard					
LCS Standard					

Sr#	SampleID	Data File Name	Date-Time	Operator	Status
1	BFB	VY016242.D	07 Nov 2023 08:03	SY/MD	Ok
2	VSTDCCC050	VY016243.D	07 Nov 2023 08:35	SY/MD	Ok,M
3	VY1107SBL01	VY016244.D	07 Nov 2023 09:31	SY/MD	Ok
4	VY1107SBS01	VY016245.D	07 Nov 2023 10:08	SY/MD	Ok,M
5	VY1107SBSD01	VY016246.D	07 Nov 2023 10:31	SY/MD	Ok,M
6	O5252-03	VY016247.D	07 Nov 2023 11:34	SY/MD	Ok
7	BLK	VY016248.D	07 Nov 2023 12:45	SY/MD	Ok
8	O5292-03	VY016249.D	07 Nov 2023 16:41	SY/MD	ReRun
9	O5291-03	VY016250.D	07 Nov 2023 17:04	SY/MD	Ok
10	VSTDCCC050	VY016251.D	07 Nov 2023 17:27	SY/MD	Ok,M

M : Manual Integration