



284 Sheffield Street, Mountainside, NJ 07092 Phone: 908 789 8900 Fax: 908 789 8922

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH Contract: RMJE02
Lab Code: CHEM Case No.: O5252 SAS No.: O5252 SDG NO.: O5252
Lab File ID: VY016243.D Date Analyzed: 11/07/2023
Instrument ID: MSVOA_Y Time Analyzed: 08:35
GC Column: RXI-624 ID: 0.25 (mm) Heated Purge: (Y/N) Y

	IS1 AREA #	RT #	IS2 AREA #	RT #	IS3 AREA #	RT #
12 HOUR STD	164275	7.80	262616	8.70	222967	11.50
UPPER LIMIT	328550	8.295	525232	9.197	445934	12.002
LOWER LIMIT	82137.5	7.295	131308	8.197	111484	11.002
EPA SAMPLE NO.						
WASTE-VOC	104317	7.80	183103	8.70	164024	11.50
VY1107SBL01	106864	7.80	189173	8.70	181795	11.50
VY1107SBS01	169940	7.80	275085	8.70	237625	11.50
VY1107SBSD01	160601	7.80	263538	8.70	230283	11.50

IS1 = Pentafluorobenzene
IS2 = 1,4-Difluorobenzene
IS3 = Chlorobenzene-d5

AREA UPPER LIMIT = +100% of internal standard area
AREA LOWER LIMIT = -50% of internal standard area
RT UPPER LIMIT = +0.50 minutes of internal standard RT
RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
* Values outside of QC limits.