



284 Sheffield Street, Mountainside, NJ 07092 Phone: 908 789 8900 Fax: 908 789 8922

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH Contract: RMJE02
Lab Code: CHEM Case No.: O5252 SAS No.: O5252 SDG NO.: O5252
Lab File ID: VY016243.D Date Analyzed: 11/07/2023
Instrument ID: MSVOA_Y Time Analyzed: 08:35
GC Column: RXI-624 ID: 0.25 (mm) Heated Purge: (Y/N) Y

	IS4 AREA #	RT #				
12 HOUR STD	100419	13.434				
UPPER LIMIT	200838	13.934				
LOWER LIMIT	50209.5	12.934				
EPA SAMPLE NO.						
WASTE-VOC	59390	13.43				
VY1107SBL01	83712	13.43				
VY1107SBS01	111499	13.43				
VY1107SBSD01	109001	13.43				

IS4 = 1,4-Dichlorobenzene-d4

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.