



284 Sheffield Street, Mountainside NJ 07092 (908)-789-8900 Fax : 908 789 8922

Continuing Calibration Report for SequenceID : FC020124AL

Parameter	AreaCount	Conc.	RT_Min	RT_Max	Response Factor	AVGRF	%DEV
-----------	-----------	-------	--------	--------	-----------------	-------	------

File ID : FC065579.D

Aliphatic C9-C12	9162902.000	60.000	3.185	6.461	152715.033	158284.239	3.518
Aliphatic C12-C16	6134915.000	40.000	6.462	9.845	153372.875	158382.836	3.163
Aliphatic C16-C21	9349041.000	60.000	9.846	13.191	155817.350	159987.165	2.606
Aliphatic C21-C28	11848213.000	80.000	13.192	16.832	148102.663	153950.508	3.799
Aliphatic C28-C40	15943271.000	120.000	16.833	21.647	132860.592	143560.801	7.453
Aliphatic EPH	52438342.000	360.000	3.185	21.647	145662.061	152708.151	4.614