



284 Sheffield Street, Mountainside NJ 07092 (908)-789-8900 Fax : 908 789 8922

Continuing Calibration Report for SequenceID : FC020124AL

Parameter	AreaCount	Conc.	RT_Min	RT_Max	Response Factor	AVGRF	%DEV
-----------	-----------	-------	--------	--------	-----------------	-------	------

File ID : FC065589.D

Aliphatic C9-C12	9338021.000	60.000	3.185	6.461	155633.683	158284.239	1.675
Aliphatic C12-C16	6126184.000	40.000	6.462	9.845	153154.600	158382.836	3.301
Aliphatic C16-C21	9113961.000	60.000	9.846	13.191	151899.350	159987.165	5.055
Aliphatic C21-C28	11703129.000	80.000	13.192	16.832	146289.113	153950.508	4.977
Aliphatic C28-C40	16482268.000	120.000	16.833	21.647	137352.233	143560.801	4.325
Aliphatic EPH	52763563.000	360.000	3.185	21.647	146565.453	152708.151	4.023