



284 Sheffield Street, Mountainside NJ 07092 (908)-789-8900 Fax : 908 789 8922

Continuing Calibration Report for SequenceID : FD020124AR

Parameter	AreaCount	Conc.	RT_Min	RT_Max	Response Factor	AVGRF	%DEV
-----------	-----------	-------	--------	--------	-----------------	-------	------

File ID : FD047144.D

Aromatic C10-C12	6209574.000	40.000	4.184	5.910	155239.350	161653.188	3.968
Aromatic C12-C16	9410937.000	60.000	5.911	8.521	156848.950	163230.209	3.909
Aromatic C16-C21	12052759.000	80.000	8.522	12.782	150659.488	156440.299	3.695
Aromatic C21-C36	24720900.000	180.000	12.783	18.171	137338.333	144232.064	4.780
Aromatic EPH	52394170.000	360.000	4.184	18.171	145539.361	152047.043	4.280