



284 Sheffield Street, Mountainside NJ 07092 (908)-789-8900 Fax : 908 789 8922

Continuing Calibration Report for SequenceID : FD020124AR

Parameter	AreaCount	Conc.	RT_Min	RT_Max	Response Factor	AVGRF	%DEV
-----------	-----------	-------	--------	--------	-----------------	-------	------

File ID : FD047154.D

Aromatic C10-C12	6407112.000	40.000	4.184	5.910	160177.800	161653.188	0.913
Aromatic C12-C16	9724710.000	60.000	5.911	8.521	162078.500	163230.209	0.706
Aromatic C16-C21	12438477.000	80.000	8.522	12.782	155480.963	156440.299	0.613
Aromatic C21-C36	25339138.000	180.000	12.783	18.171	140772.989	144232.064	2.398
Aromatic EPH	53909437.000	360.000	4.184	18.171	149748.436	152047.043	1.512