



284 Sheffield Street, Mountainside NJ 07092 (908)-789-8900 Fax : 908 789 8922

Continuing Calibration Report for SequenceID : FE020624AL

Parameter	AreaCount	Conc.	RT_Min	RT_Max	Response Factor	AVGRF	%DEV
-----------	-----------	-------	--------	--------	-----------------	-------	------

File ID : FE046594.D

Aliphatic C9-C12	9483420.000	60.000	3.103	6.721	158057.000	163608.101	3.393
Aliphatic C12-C16	6272023.000	40.000	6.722	10.157	156800.575	167387.982	6.325
Aliphatic C16-C21	8974221.000	60.000	10.158	13.524	149570.350	170304.781	12.175
Aliphatic C21-C28	11091950.000	80.000	13.525	17.190	138649.375	162263.126	14.553
Aliphatic C28-C40	15358642.000	120.000	17.191	22.057	127988.683	140116.521	8.656
Aliphatic EPH	51180256.000	360.000	3.103	22.057	142167.378	157014.791	9.456