



284 Sheffield Street, Mountainside NJ 07092 (908)-789-8900 Fax : 908 789 8922

Continuing Calibration Report for SequenceID : FE020624AL

Parameter	AreaCount	Conc.	RT_Min	RT_Max	Response Factor	AVGRF	%DEV
-----------	-----------	-------	--------	--------	-----------------	-------	------

File ID : FE046604.D

Aliphatic C9-C12	9379572.000	60.000	3.103	6.721	156326.200	163608.101	4.451
Aliphatic C12-C16	6225261.000	40.000	6.722	10.157	155631.525	167387.982	7.023
Aliphatic C16-C21	8933460.000	60.000	10.158	13.524	148891.000	170304.781	12.574
Aliphatic C21-C28	11098916.000	80.000	13.525	17.190	138736.450	162263.126	14.499
Aliphatic C28-C40	15317788.000	120.000	17.191	22.057	127648.233	140116.521	8.899
Aliphatic EPH	50954997.000	360.000	3.103	22.057	141541.658	157014.791	9.855