



284 Sheffield Street, Mountainside NJ 07092 (908)-789-8900 Fax : 908 789 8922

Continuing Calibration Report for SequenceID : FE020624AL

Parameter	AreaCount	Conc.	RT_Min	RT_Max	Response Factor	AVGRF	%DEV
-----------	-----------	-------	--------	--------	-----------------	-------	------

File ID : FE046615.D

Aliphatic C9-C12	9293668.000	60.000	3.103	6.721	154894.467	163608.101	5.326
Aliphatic C12-C16	6198118.000	40.000	6.722	10.157	154952.950	167387.982	7.429
Aliphatic C16-C21	8968541.000	60.000	10.158	13.524	149475.683	170304.781	12.230
Aliphatic C21-C28	11179045.000	80.000	13.525	17.190	139738.063	162263.126	13.882
Aliphatic C28-C40	15502908.000	120.000	17.191	22.057	129190.900	140116.521	7.798
Aliphatic EPH	51142280.000	360.000	3.103	22.057	142061.889	157014.791	9.523