



284 Sheffield Street, Mountainside NJ 07092 (908)-789-8900 Fax : 908 789 8922

Continuing Calibration Report for SequenceID : FE020624AL

Parameter	AreaCount	Conc.	RT_Min	RT_Max	Response Factor	AVGRF	%DEV
-----------	-----------	-------	--------	--------	-----------------	-------	------

File ID : FE046627.D

Aliphatic C9-C12	9247980.000	60.000	3.103	6.721	154133.000	163608.101	5.791
Aliphatic C12-C16	6166223.000	40.000	6.722	10.157	154155.575	167387.982	7.905
Aliphatic C16-C21	8892818.000	60.000	10.158	13.524	148213.633	170304.781	12.972
Aliphatic C21-C28	11149281.000	80.000	13.525	17.190	139366.013	162263.126	14.111
Aliphatic C28-C40	15843563.000	120.000	17.191	22.057	132029.692	140116.521	5.772
Aliphatic EPH	51299865.000	360.000	3.103	22.057	142499.625	157014.791	9.244