



284 Sheffield Street, Mountainside NJ 07092 (908)-789-8900 Fax : 908 789 8922

Continuing Calibration Report for SequenceID : FG021324AL

Parameter	AreaCount	Conc.	RT_Min	RT_Max	Response Factor	AVGRF	%DEV
-----------	-----------	-------	--------	--------	-----------------	-------	------

File ID : FG013499.D

Aliphatic C9-C12	7946510.000	60.000	3.229	6.851	132441.833	131827.767	-0.466
Aliphatic C12-C16	5444729.000	40.000	6.852	10.298	136118.225	135291.235	-0.611
Aliphatic C16-C21	8048455.000	60.000	10.299	13.676	134140.917	133403.467	-0.553
Aliphatic C21-C28	9960640.000	80.000	13.677	17.352	124508.000	123480.725	-0.832
Aliphatic C28-C40	12916850.000	120.000	17.353	22.308	107640.417	108116.277	0.440
Aliphatic EPH	44317184.000	360.000	3.229	22.308	123103.289	122716.485	-0.315