



284 Sheffield Street, Mountainside NJ 07092 (908)-789-8900 Fax : 908 789 8922

Continuing Calibration Report for SequenceID : FG021324AL

Parameter	AreaCount	Conc.	RT_Min	RT_Max	Response Factor	AVGRF	%DEV
-----------	-----------	-------	--------	--------	-----------------	-------	------

File ID : FG013509.D

Aliphatic C9-C12	8006740.000	60.000	3.229	6.851	133445.667	131827.767	-1.227
Aliphatic C12-C16	5451849.000	40.000	6.852	10.298	136296.225	135291.235	-0.743
Aliphatic C16-C21	8125780.000	60.000	10.299	13.676	135429.667	133403.467	-1.519
Aliphatic C21-C28	10170558.000	80.000	13.677	17.352	127131.975	123480.725	-2.957
Aliphatic C28-C40	13539396.000	120.000	17.353	22.308	112828.300	108116.277	-4.358
Aliphatic EPH	45294323.000	360.000	3.229	22.308	125817.564	122716.485	-2.527