



284 Sheffield Street, Mountainside, NJ 07092 Phone: 908 789 8900 Fax: 908 789 8922

CALIBRATION VERIFICATION SUMMARY

Contract: CHEM02

Lab Code: CHEM Case No.: P1187 SAS No.: P1187 SDG NO.: P1187

Continuing Calib Date: 01/23/2024 Initial Calibration Date(s): 01/11/2024 01/11/2024

Continuing Calib Time: 03:11 Initial Calibration Time(s): 14:35 16:23

GC Column: RTX-CLP ID: 0.32 (mm)

COMPOUND	CCAL RT	AVG RT	RT WINDOW FROM TO		DIFF RT
DICAMBA	6.49	6.49	6.39	6.59	0.00
2,4-DCAA	6.35	6.35	6.25	6.45	0.00
DICHLORPROP	7.07	7.08	6.98	7.18	0.01
2,4-D	7.27	7.28	7.18	7.38	0.01
2,4,5-TP(Silvex)	8.02	8.02	7.92	8.12	0.00
2,4,5-T	8.28	8.28	8.18	8.38	0.00
2,4-DB	8.78	8.79	8.69	8.89	0.01
Dinoseb	9.83	9.83	9.73	9.93	0.00



284 Sheffield Street, Mountainside, NJ 07092 Phone: 908 789 8900 Fax: 908 789 8922

CALIBRATION VERIFICATION SUMMARY

Contract: CHEM02

Lab Code: CHEM Case No.: P1187 SAS No.: P1187 SDG NO.: P1187

Continuing Calib Date: 01/23/2024 Initial Calibration Date(s): 01/11/2024 01/11/2024

Continuing Calib Time: 03:11 Initial Calibration Time(s): 14:35 16:23

GC Column: RTX-CLP2 ID: 0.32 (mm)

COMPOUND	CCAL RT	AVG RT	RT WINDOW FROM TO		DIFF RT
DICAMBA	6.92	6.92	6.82	7.02	0.00
2,4-DCAA	6.76	6.76	6.66	6.86	0.00
DICHLORPROP	7.53	7.53	7.43	7.63	0.00
2,4-D	7.81	7.82	7.72	7.92	0.01
2,4,5-TP(Silvex)	8.59	8.60	8.50	8.70	0.01
2,4,5-T	8.96	8.96	8.86	9.06	0.00
2,4-DB	9.46	9.47	9.37	9.57	0.01
Dinoseb	9.79	9.79	9.69	9.89	0.00