



284 Sheffield Street, Mountainside, NJ 07092 Phone: 908 789 8900 Fax: 908 789 8922

CALIBRATION VERIFICATION SUMMARY

Contract: CHEM02

Lab Code: CHEM Case No.: P1187 SAS No.: P1187 SDG NO.: P1187

GC Column: RTX-CLP ID: 0.32 (mm) Initi. Calib. Date(s): 01/11/2024 01/11/2024

Client Sample No.: CCAL01 Date Analyzed: 01/22/2024

Lab Sample No.: HSTDCCC750 Data File : PS025605.D Time Analyzed: 19:58

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
2,4,5-T	8.277	8.181	8.381	660.070	712.500	-7.4
2,4,5-TP(Silvex)	8.020	7.923	8.123	714.090	712.500	0.2
2,4-D	7.273	7.176	7.376	664.540	705.000	-5.7
2,4-DB	8.783	8.686	8.886	728.620	712.500	2.3
2,4-DCAA	6.349	6.250	6.450	735.310	750.000	-2.0
DICAMBA	6.489	6.391	6.591	703.890	705.000	-0.2
DICHLORPROP	7.074	6.977	7.177	693.500	705.000	-1.6
Dinoseb	9.829	9.733	9.933	715.990	705.000	1.6



284 Sheffield Street, Mountainside, NJ 07092 Phone: 908 789 8900 Fax: 908 789 8922

CALIBRATION VERIFICATION SUMMARY

Contract: CHEM02

Lab Code: CHEM Case No.: P1187 SAS No.: P1187 SDG NO.: P1187

GC Column: RTX-CLP2 ID: 0.32 (mm) Initi. Calib. Date(s): 01/11/2024 01/11/2024

Client Sample No.: CCAL01 Date Analyzed: 01/22/2024

Lab Sample No.: HSTDCCC750 Data File : PS025605.D Time Analyzed: 19:58

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
2,4,5-T	8.961	8.864	9.064	737.330	712.500	3.5
2,4,5-TP(Silvex)	8.591	8.495	8.695	725.630	712.500	1.8
2,4-D	7.812	7.715	7.915	706.430	705.000	0.2
2,4-DB	9.464	9.368	9.568	724.840	712.500	1.7
2,4-DCAA	6.758	6.661	6.861	741.000	750.000	-1.2
DICAMBA	6.917	6.820	7.020	711.770	705.000	1.0
DICHLORPROP	7.528	7.432	7.632	694.350	705.000	-1.5
Dinoseb	9.790	9.693	9.893	737.300	705.000	4.6