



284 Sheffield Street, Mountainside, NJ 07092 Phone: 908 789 8900 Fax: 908 789 8922

CALIBRATION VERIFICATION SUMMARY

Contract: CHEM02

Lab Code: CHEM Case No.: P1187 SAS No.: P1187 SDG NO.: P1187

GC Column: RTX-CLP ID: 0.32 (mm) Initi. Calib. Date(s): 01/11/2024 01/11/2024

Client Sample No.: CCAL02 Date Analyzed: 01/22/2024

Lab Sample No.: HSTDCCC750 Data File : PS025613.D Time Analyzed: 23:10

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
2,4,5-T	8.277	8.181	8.381	681.660	712.500	-4.3
2,4,5-TP(Silvex)	8.020	7.923	8.123	677.300	712.500	-4.9
2,4-D	7.273	7.176	7.376	644.240	705.000	-8.6
2,4-DB	8.782	8.686	8.886	720.040	712.500	1.1
2,4-DCAA	6.349	6.250	6.450	701.010	750.000	-6.5
DICAMBA	6.489	6.391	6.591	671.360	705.000	-4.8
DICHLORPROP	7.074	6.977	7.177	666.230	705.000	-5.5
Dinoseb	9.829	9.733	9.933	688.360	705.000	-2.4



284 Sheffield Street, Mountainside, NJ 07092 Phone: 908 789 8900 Fax: 908 789 8922

CALIBRATION VERIFICATION SUMMARY

Contract: CHEM02

Lab Code: CHEM Case No.: P1187 SAS No.: P1187 SDG NO.: P1187

GC Column: RTX-CLP2 ID: 0.32 (mm) Initi. Calib. Date(s): 01/11/2024 01/11/2024

Client Sample No.: CCAL02 Date Analyzed: 01/22/2024

Lab Sample No.: HSTDCCC750 Data File : PS025613.D Time Analyzed: 23:10

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
2,4,5-T	8.960	8.864	9.064	719.910	712.500	1.0
2,4,5-TP(Silvex)	8.590	8.495	8.695	701.310	712.500	-1.6
2,4-D	7.811	7.715	7.915	683.330	705.000	-3.1
2,4-DB	9.463	9.368	9.568	719.760	712.500	1.0
2,4-DCAA	6.758	6.661	6.861	714.170	750.000	-4.8
DICAMBA	6.916	6.820	7.020	681.660	705.000	-3.3
DICHLORPROP	7.528	7.432	7.632	667.230	705.000	-5.4
Dinoseb	9.789	9.693	9.893	717.310	705.000	1.7