



284 Sheffield Street, Mountainside, NJ 07092 Phone: 908 789 8900 Fax: 908 789 8922

CALIBRATION VERIFICATION SUMMARY

Contract: CHEM02

Lab Code: CHEM Case No.: P1187 SAS No.: P1187 SDG NO.: P1187

GC Column: RTX-CLP ID: 0.32 (mm) Initi. Calib. Date(s): 01/11/2024 01/11/2024

Client Sample No.: CCAL03 Date Analyzed: 01/23/2024

Lab Sample No.: HSTDCCC750 Data File : PS025622.D Time Analyzed: 03:11

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
2,4,5-T	8.277	8.181	8.381	633.850	712.500	-11.0
2,4,5-TP(Silvex)	8.019	7.923	8.123	664.870	712.500	-6.7
2,4-D	7.273	7.176	7.376	633.410	705.000	-10.2
2,4-DB	8.783	8.686	8.886	692.800	712.500	-2.8
2,4-DCAA	6.349	6.250	6.450	696.660	750.000	-7.1
DICAMBA	6.489	6.391	6.591	665.480	705.000	-5.6
DICHLORPROP	7.074	6.977	7.177	660.090	705.000	-6.4
Dinoseb	9.829	9.733	9.933	680.180	705.000	-3.5



284 Sheffield Street, Mountainside, NJ 07092 Phone: 908 789 8900 Fax: 908 789 8922

CALIBRATION VERIFICATION SUMMARY

Contract: CHEM02

Lab Code: CHEM Case No.: P1187 SAS No.: P1187 SDG NO.: P1187

GC Column: RTX-CLP2 ID: 0.32 (mm) Initi. Calib. Date(s): 01/11/2024 01/11/2024

Client Sample No.: CCAL03 Date Analyzed: 01/23/2024

Lab Sample No.: HSTDCCC750 Data File : PS025622.D Time Analyzed: 03:11

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
2,4,5-T	8.960	8.864	9.064	705.830	712.500	-0.9
2,4,5-TP(Silvex)	8.591	8.495	8.695	692.340	712.500	-2.8
2,4-D	7.811	7.715	7.915	670.470	705.000	-4.9
2,4-DB	9.464	9.368	9.568	703.360	712.500	-1.3
2,4-DCAA	6.758	6.661	6.861	704.250	750.000	-6.1
DICAMBA	6.916	6.820	7.020	673.860	705.000	-4.4
DICHLORPROP	7.528	7.432	7.632	661.240	705.000	-6.2
Dinoseb	9.788	9.693	9.893	700.990	705.000	-0.6