



284 Sheffield Street, Mountainside, NJ 07092 Phone: 908 789 8900 Fax: 908 789 8922

CALIBRATION VERIFICATION SUMMARY

Contract: CHEM02

Lab Code: CHEM Case No.: P1187 SAS No.: P1187 SDG NO.: P1187

Continuing Calib Date: 01/25/2024 Initial Calibration Date(s): 01/23/2024 01/24/2024

Continuing Calib Time: 12:07 Initial Calibration Time(s): 17:32 01:03

GC Column: ZB-MR1 ID: 0.32 (mm)

COMPOUND	CCAL RT	AVG RT	RT WINDOW		DIFF RT
			FROM	TO	
Aroclor-1254-1 (1)	6.36	6.36	6.26	6.46	0.00
Aroclor-1254-2 (2)	6.58	6.58	6.48	6.68	0.01
Aroclor-1254-3 (3)	6.94	6.94	6.84	7.04	0.00
Aroclor-1254-4 (4)	7.23	7.23	7.13	7.33	0.00
Aroclor-1254-5 (5)	7.65	7.65	7.55	7.75	0.00
Tetrachloro-m-xylene	4.33	4.33	4.23	4.43	0.00
Decachlorobiphenyl	10.06	10.06	9.96	10.16	0.00



284 Sheffield Street, Mountainside, NJ 07092 Phone: 908 789 8900 Fax: 908 789 8922

CALIBRATION VERIFICATION SUMMARY

Contract: CHEM02

Lab Code: CHEM Case No.: P1187 SAS No.: P1187 SDG NO.: P1187

Continuing Calib Date: 01/25/2024 Initial Calibration Date(s): 01/23/2024 01/24/2024

Continuing Calib Time: 12:07 Initial Calibration Time(s): 17:32 01:03

GC Column: ZB-MR2 ID: 0.32 (mm)

COMPOUND	CCAL RT	AVG RT	RT WINDOW		DIFF RT
			FROM	TO	
Aroclor-1254-1 (1)	5.53	5.53	5.43	5.63	0.00
Aroclor-1254-2 (2)	5.68	5.68	5.58	5.78	0.00
Aroclor-1254-3 (3)	6.09	6.09	5.99	6.19	0.00
Aroclor-1254-4 (4)	6.32	6.32	6.22	6.42	0.00
Aroclor-1254-5 (5)	6.74	6.74	6.64	6.84	0.00
Tetrachloro-m-xylene	3.63	3.63	3.53	3.73	0.00
Decachlorobiphenyl	8.68	8.68	8.58	8.78	0.00