



284 Sheffield Street, Mountainside, NJ 07092 Phone: 908 789 8900 Fax: 908 789 8922

### CALIBRATION VERIFICATION SUMMARY

Contract: CHEM02

Lab Code: CHEM Case No.: P1187 SAS No.: P1187 SDG NO.: P1187

Continuing Calib Date: 01/26/2024 Initial Calibration Date(s): 01/23/2024 01/24/2024

Continuing Calib Time: 13:09 Initial Calibration Time(s): 17:32 01:03

GC Column: ZB-MR1 ID: 0.32 (mm)

| COMPOUND             |     | CCAL<br>RT | AVG<br>RT | RT WINDOW |       | DIFF<br>RT |
|----------------------|-----|------------|-----------|-----------|-------|------------|
|                      |     |            |           | FROM      | TO    |            |
| Aroclor-1016-1       | (1) | 5.50       | 5.50      | 5.40      | 5.60  | 0.00       |
| Aroclor-1016-2       | (2) | 5.52       | 5.52      | 5.42      | 5.62  | 0.00       |
| Aroclor-1016-3       | (3) | 5.59       | 5.59      | 5.49      | 5.69  | 0.00       |
| Aroclor-1016-4       | (4) | 5.68       | 5.68      | 5.58      | 5.78  | 0.00       |
| Aroclor-1016-5       | (5) | 5.98       | 5.98      | 5.88      | 6.08  | 0.00       |
| Aroclor-1260-1       | (1) | 7.11       | 7.11      | 7.01      | 7.21  | 0.00       |
| Aroclor-1260-2       | (2) | 7.37       | 7.37      | 7.27      | 7.47  | 0.01       |
| Aroclor-1260-3       | (3) | 7.73       | 7.73      | 7.63      | 7.83  | 0.00       |
| Aroclor-1260-4       | (4) | 7.95       | 7.95      | 7.85      | 8.05  | 0.00       |
| Aroclor-1260-5       | (5) | 8.27       | 8.27      | 8.17      | 8.37  | 0.00       |
| Tetrachloro-m-xylene |     | 4.33       | 4.33      | 4.23      | 4.43  | 0.00       |
| Decachlorobiphenyl   |     | 10.06      | 10.06     | 9.96      | 10.16 | 0.00       |



284 Sheffield Street, Mountainside, NJ 07092 Phone: 908 789 8900 Fax: 908 789 8922

### CALIBRATION VERIFICATION SUMMARY

Contract: CHEM02

Lab Code: CHEM Case No.: P1187 SAS No.: P1187 SDG NO.: P1187

Continuing Calib Date: 01/26/2024 Initial Calibration Date(s): 01/23/2024 01/24/2024

Continuing Calib Time: 13:09 Initial Calibration Time(s): 17:32 01:03

GC Column: ZB-MR2 ID: 0.32 (mm)

| COMPOUND             | CCAL<br>RT | AVG<br>RT | RT WINDOW |      | DIFF<br>RT |
|----------------------|------------|-----------|-----------|------|------------|
|                      |            |           | FROM      | TO   |            |
| Aroclor-1016-1 (1)   | 4.72       | 4.72      | 4.62      | 4.82 | 0.00       |
| Aroclor-1016-2 (2)   | 4.74       | 4.74      | 4.64      | 4.84 | 0.00       |
| Aroclor-1016-3 (3)   | 4.92       | 4.92      | 4.82      | 5.02 | 0.00       |
| Aroclor-1016-4 (4)   | 4.96       | 4.96      | 4.86      | 5.06 | 0.00       |
| Aroclor-1016-5 (5)   | 5.18       | 5.18      | 5.08      | 5.28 | 0.00       |
| Aroclor-1260-1 (1)   | 6.22       | 6.22      | 6.12      | 6.32 | 0.00       |
| Aroclor-1260-2 (2)   | 6.41       | 6.41      | 6.31      | 6.51 | 0.00       |
| Aroclor-1260-3 (3)   | 6.56       | 6.56      | 6.46      | 6.66 | 0.00       |
| Aroclor-1260-4 (4)   | 7.04       | 7.04      | 6.94      | 7.14 | 0.01       |
| Aroclor-1260-5 (5)   | 7.28       | 7.28      | 7.18      | 7.38 | 0.00       |
| Tetrachloro-m-xylene | 3.63       | 3.63      | 3.53      | 3.73 | 0.00       |
| Decachlorobiphenyl   | 8.68       | 8.68      | 8.58      | 8.78 | 0.00       |