



284 Sheffield Street, Mountainside, NJ 07092 Phone: 908 789 8900 Fax: 908 789 8922

CALIBRATION VERIFICATION SUMMARY

Contract: CHEM02

Lab Code: CHEM Case No.: P1187 SAS No.: P1187 SDG NO.: P1187

Continuing Calib Date: 01/27/2024 Initial Calibration Date(s): 01/23/2024 01/24/2024

Continuing Calib Time: 04:47 Initial Calibration Time(s): 17:32 01:03

GC Column: ZB-MR1 ID: 0.32 (mm)

COMPOUND		CCAL RT	AVG RT	RT WINDOW		DIFF RT
				FROM	TO	
Aroclor-1016-1	(1)	5.50	5.50	5.40	5.60	0.00
Aroclor-1016-2	(2)	5.52	5.52	5.42	5.62	0.00
Aroclor-1016-3	(3)	5.58	5.59	5.49	5.69	0.01
Aroclor-1016-4	(4)	5.68	5.68	5.58	5.78	0.00
Aroclor-1016-5	(5)	5.98	5.98	5.88	6.08	0.00
Aroclor-1260-1	(1)	7.11	7.11	7.01	7.21	0.00
Aroclor-1260-2	(2)	7.36	7.37	7.27	7.47	0.01
Aroclor-1260-3	(3)	7.73	7.73	7.63	7.83	0.01
Aroclor-1260-4	(4)	7.95	7.95	7.85	8.05	0.00
Aroclor-1260-5	(5)	8.27	8.27	8.17	8.37	0.00
Tetrachloro-m-xylene		4.33	4.33	4.23	4.43	0.00
Decachlorobiphenyl		10.06	10.06	9.96	10.16	0.00



284 Sheffield Street, Mountainside, NJ 07092 Phone: 908 789 8900 Fax: 908 789 8922

CALIBRATION VERIFICATION SUMMARY

Contract: CHEM02

Lab Code: CHEM Case No.: P1187 SAS No.: P1187 SDG NO.: P1187

Continuing Calib Date: 01/27/2024 Initial Calibration Date(s): 01/23/2024 01/24/2024

Continuing Calib Time: 04:47 Initial Calibration Time(s): 17:32 01:03

GC Column: ZB-MR2 ID: 0.32 (mm)

COMPOUND	CCAL RT	AVG RT	RT WINDOW		DIFF RT
			FROM	TO	
Aroclor-1016-1 (1)	4.72	4.72	4.62	4.82	0.00
Aroclor-1016-2 (2)	4.74	4.74	4.64	4.84	0.00
Aroclor-1016-3 (3)	4.92	4.92	4.82	5.02	0.00
Aroclor-1016-4 (4)	4.96	4.96	4.86	5.06	0.00
Aroclor-1016-5 (5)	5.18	5.18	5.08	5.28	0.00
Aroclor-1260-1 (1)	6.22	6.22	6.12	6.32	0.00
Aroclor-1260-2 (2)	6.40	6.41	6.31	6.51	0.01
Aroclor-1260-3 (3)	6.56	6.56	6.46	6.66	0.00
Aroclor-1260-4 (4)	7.03	7.04	6.94	7.14	0.01
Aroclor-1260-5 (5)	7.28	7.28	7.18	7.38	0.00
Tetrachloro-m-xylene	3.63	3.63	3.53	3.73	0.00
Decachlorobiphenyl	8.67	8.68	8.58	8.78	0.01