



284 Sheffield Street, Mountainside, NJ 07092 Phone: 908 789 8900 Fax: 908 789 8922

CALIBRATION VERIFICATION SUMMARY

Contract: CHEM02

Lab Code: CHEM Case No.: P1187 SAS No.: P1187 SDG NO.: P1187

Continuing Calib Date: 02/02/2024 Initial Calibration Date(s): 02/01/2024 02/01/2024

Continuing Calib Time: 10:05 Initial Calibration Time(s): 10:33 13:32

GC Column: ZB-MR1 ID: 0.32 (mm)

COMPOUND		CCAL RT	AVG RT	RT WINDOW		DIFF RT
				FROM	TO	
Aroclor-1016-1	(1)	5.67	5.67	5.57	5.77	0.00
Aroclor-1016-2	(2)	5.70	5.70	5.60	5.80	0.00
Aroclor-1016-3	(3)	5.76	5.76	5.66	5.86	0.00
Aroclor-1016-4	(4)	5.86	5.86	5.76	5.96	0.00
Aroclor-1016-5	(5)	6.16	6.16	6.06	6.26	0.00
Aroclor-1260-1	(1)	7.29	7.29	7.19	7.39	0.00
Aroclor-1260-2	(2)	7.55	7.55	7.45	7.65	0.00
Aroclor-1260-3	(3)	7.91	7.91	7.81	8.01	0.00
Aroclor-1260-4	(4)	8.14	8.14	8.04	8.24	0.00
Aroclor-1260-5	(5)	8.47	8.47	8.37	8.57	0.00
Tetrachloro-m-xylene		4.49	4.49	4.39	4.59	0.00
Decachlorobiphenyl		10.36	10.36	10.26	10.46	0.00



284 Sheffield Street, Mountainside, NJ 07092 Phone: 908 789 8900 Fax: 908 789 8922

CALIBRATION VERIFICATION SUMMARY

Contract: CHEM02

Lab Code: CHEM Case No.: P1187 SAS No.: P1187 SDG NO.: P1187

Continuing Calib Date: 02/02/2024 Initial Calibration Date(s): 02/01/2024 02/01/2024

Continuing Calib Time: 10:05 Initial Calibration Time(s): 10:33 13:32

GC Column: ZB-MR2 ID: 0.32 (mm)

COMPOUND	CCAL RT	AVG RT	RT WINDOW		DIFF RT
			FROM	TO	
Aroclor-1016-1 (1)	4.74	4.74	4.64	4.84	0.00
Aroclor-1016-2 (2)	4.76	4.76	4.66	4.86	0.00
Aroclor-1016-3 (3)	4.94	4.94	4.84	5.04	0.00
Aroclor-1016-4 (4)	4.98	4.98	4.88	5.08	0.00
Aroclor-1016-5 (5)	5.19	5.20	5.10	5.30	0.01
Aroclor-1260-1 (1)	6.23	6.24	6.14	6.34	0.01
Aroclor-1260-2 (2)	6.42	6.43	6.33	6.53	0.01
Aroclor-1260-3 (3)	6.58	6.58	6.48	6.68	0.00
Aroclor-1260-4 (4)	7.05	7.06	6.96	7.16	0.01
Aroclor-1260-5 (5)	7.30	7.30	7.20	7.40	0.00
Tetrachloro-m-xylene	3.65	3.65	3.55	3.75	0.00
Decachlorobiphenyl	8.70	8.70	8.60	8.80	0.00