



284 Sheffield Street, Mountainside, NJ 07092 Phone: 908 789 8900 Fax: 908 789 8922

CALIBRATION VERIFICATION SUMMARY

Contract: CHEM02

Lab Code: CHEM Case No.: P1187 SAS No.: P1187 SDG NO.: P1187

Continuing Calib Date: 01/24/2024 Initial Calibration Date(s): 01/17/2024 01/17/2024

Continuing Calib Time: 12:01 Initial Calibration Time(s): 09:23 16:56

GC Column: ZB-MR1 ID: 0.32 (mm)

COMPOUND		CCAL RT	AVG RT	RT WINDOW		DIFF RT
				FROM	TO	
Aroclor-1016-1	(1)	4.60	4.59	4.49	4.69	0.00
Aroclor-1016-2	(2)	4.62	4.61	4.51	4.71	0.00
Aroclor-1016-3	(3)	4.67	4.67	4.57	4.77	0.00
Aroclor-1016-4	(4)	4.77	4.77	4.67	4.87	0.00
Aroclor-1016-5	(5)	5.06	5.06	4.96	5.16	0.00
Aroclor-1260-1	(1)	6.18	6.18	6.08	6.28	0.00
Aroclor-1260-2	(2)	6.45	6.45	6.35	6.55	0.00
Aroclor-1260-3	(3)	6.81	6.81	6.71	6.91	0.00
Aroclor-1260-4	(4)	7.03	7.03	6.93	7.13	0.00
Aroclor-1260-5	(5)	7.35	7.35	7.25	7.45	0.00
Tetrachloro-m-xylene		3.47	3.47	3.37	3.57	0.00
Decachlorobiphenyl		8.78	8.78	8.68	8.88	0.00



284 Sheffield Street, Mountainside, NJ 07092 Phone: 908 789 8900 Fax: 908 789 8922

CALIBRATION VERIFICATION SUMMARY

Contract: CHEM02

Lab Code: CHEM Case No.: P1187 SAS No.: P1187 SDG NO.: P1187

Continuing Calib Date: 01/24/2024 Initial Calibration Date(s): 01/17/2024 01/17/2024

Continuing Calib Time: 12:01 Initial Calibration Time(s): 09:23 16:56

GC Column: ZB-MR2 ID: 0.32 (mm)

COMPOUND	CCAL RT	AVG RT	RT WINDOW		DIFF RT
			FROM	TO	
Aroclor-1016-1 (1)	3.74	3.75	3.65	3.85	0.01
Aroclor-1016-2 (2)	3.76	3.76	3.66	3.86	0.00
Aroclor-1016-3 (3)	3.92	3.92	3.82	4.02	0.00
Aroclor-1016-4 (4)	3.97	3.97	3.87	4.07	0.00
Aroclor-1016-5 (5)	4.16	4.17	4.07	4.27	0.01
Aroclor-1260-1 (1)	5.16	5.16	5.06	5.26	0.00
Aroclor-1260-2 (2)	5.35	5.35	5.25	5.45	0.00
Aroclor-1260-3 (3)	5.49	5.50	5.40	5.60	0.01
Aroclor-1260-4 (4)	5.95	5.96	5.86	6.06	0.01
Aroclor-1260-5 (5)	6.20	6.21	6.11	6.31	0.01
Tetrachloro-m-xylene	2.74	2.75	2.65	2.85	0.01
Decachlorobiphenyl	7.52	7.53	7.43	7.63	0.01