



284 Sheffield Street, Mountainside, NJ 07092 Phone: 908 789 8900 Fax: 908 789 8922

CALIBRATION VERIFICATION SUMMARY

Contract: CHEM02

Lab Code: CHEM Case No.: P1187 SAS No.: P1187 SDG NO.: P1187

Continuing Calib Date: 01/26/2024 Initial Calibration Date(s): 01/19/2024 01/19/2024

Continuing Calib Time: 09:28 Initial Calibration Time(s): 11:22 18:02

GC Column: ZB-MR1 ID: 0.32 (mm)

COMPOUND	CCAL RT	AVG RT	RT WINDOW		DIFF RT
			FROM	TO	
Aroclor-1016-1 (1)	4.94	4.94	4.84	5.04	0.00
Aroclor-1016-2 (2)	4.97	4.97	4.87	5.07	0.00
Aroclor-1016-3 (3)	5.03	5.03	4.93	5.13	0.00
Aroclor-1016-4 (4)	5.14	5.14	5.04	5.24	0.00
Aroclor-1016-5 (5)	5.46	5.46	5.36	5.56	0.00
Aroclor-1260-1 (1)	6.69	6.69	6.59	6.79	0.00
Aroclor-1260-2 (2)	6.98	6.97	6.87	7.07	-0.01
Aroclor-1260-3 (3)	7.37	7.37	7.27	7.47	0.00
Aroclor-1260-4 (4)	7.62	7.62	7.52	7.72	0.00
Aroclor-1260-5 (5)	7.96	7.96	7.86	8.06	0.00
Tetrachloro-m-xylene	3.70	3.70	3.60	3.80	0.00
Decachlorobiphenyl	9.75	9.74	9.64	9.84	0.00



284 Sheffield Street, Mountainside, NJ 07092 Phone: 908 789 8900 Fax: 908 789 8922

CALIBRATION VERIFICATION SUMMARY

Contract: CHEM02

Lab Code: CHEM Case No.: P1187 SAS No.: P1187 SDG NO.: P1187

Continuing Calib Date: 01/26/2024 Initial Calibration Date(s): 01/19/2024 01/19/2024

Continuing Calib Time: 09:28 Initial Calibration Time(s): 11:22 18:02

GC Column: ZB-MR2 ID: 0.32 (mm)

COMPOUND		CCAL RT	AVG RT	RT WINDOW		DIFF RT
				FROM	TO	
Aroclor-1016-1	(1)	4.15	4.15	4.05	4.25	0.00
Aroclor-1016-2	(2)	4.17	4.17	4.07	4.27	0.00
Aroclor-1016-3	(3)	4.35	4.35	4.25	4.45	0.00
Aroclor-1016-4	(4)	4.40	4.41	4.31	4.51	0.01
Aroclor-1016-5	(5)	4.62	4.63	4.53	4.73	0.01
Aroclor-1260-1	(1)	5.74	5.74	5.64	5.84	0.00
Aroclor-1260-2	(2)	5.95	5.95	5.85	6.05	0.00
Aroclor-1260-3	(3)	6.11	6.11	6.01	6.21	0.00
Aroclor-1260-4	(4)	6.62	6.63	6.53	6.73	0.01
Aroclor-1260-5	(5)	6.89	6.90	6.80	7.00	0.01
Tetrachloro-m-xylene		3.00	3.00	2.90	3.10	0.00
Decachlorobiphenyl		8.37	8.38	8.28	8.48	0.01