



284 Sheffield Street, Mountainside, NJ 07092 Phone: 908 789 8900 Fax: 908 789 8922

CALIBRATION VERIFICATION SUMMARY

Contract: CHEM02

Lab Code: CHEM Case No.: P1187 SAS No.: P1187 SDG NO.: P1187

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 01/23/2024 01/23/2024

Client Sample No.: CCAL01 Date Analyzed: 01/25/2024

Lab Sample No.: AR1660CCC500 Data File : PO101357.D Time Analyzed: 15:44

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	5.500	5.400	5.600	569.020	500.000	13.8
Aroclor-1016-2	5.523	5.422	5.622	575.640	500.000	15.1
Aroclor-1016-3	5.584	5.485	5.685	576.660	500.000	15.3
Aroclor-1016-4	5.683	5.583	5.783	569.850	500.000	14.0
Aroclor-1016-5	5.978	5.879	6.079	510.090	500.000	2.0
Aroclor-1260-1	7.108	7.007	7.207	484.420	500.000	-3.1
Aroclor-1260-2	7.365	7.265	7.465	522.220	500.000	4.4
Aroclor-1260-3	7.726	7.626	7.826	490.290	500.000	-1.9
Aroclor-1260-4	7.953	7.852	8.052	521.990	500.000	4.4
Aroclor-1260-5	8.269	8.168	8.368	530.070	500.000	6.0
Decachlorobiphenyl	10.059	9.957	10.157	49.500	50.000	-1.0
Tetrachloro-m-xylene	4.333	4.234	4.434	53.630	50.000	7.3



284 Sheffield Street, Mountainside, NJ 07092 Phone: 908 789 8900 Fax: 908 789 8922

CALIBRATION VERIFICATION SUMMARY

Contract: CHEM02

Lab Code: CHEM Case No.: P1187 SAS No.: P1187 SDG NO.: P1187

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 01/23/2024 01/23/2024

Client Sample No.: CCAL01 Date Analyzed: 01/25/2024

Lab Sample No.: AR1660CCC500 Data File : PO101357.D Time Analyzed: 15:44

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.724	4.623	4.823	543.190	500.000	8.6
Aroclor-1016-2	4.743	4.642	4.842	562.680	500.000	12.5
Aroclor-1016-3	4.920	4.821	5.021	541.770	500.000	8.4
Aroclor-1016-4	4.962	4.862	5.062	501.960	500.000	0.4
Aroclor-1016-5	5.178	5.077	5.277	509.050	500.000	1.8
Aroclor-1260-1	6.218	6.117	6.317	484.110	500.000	-3.2
Aroclor-1260-2	6.407	6.306	6.506	505.140	500.000	1.0
Aroclor-1260-3	6.562	6.461	6.661	490.050	500.000	-2.0
Aroclor-1260-4	7.036	6.935	7.135	487.440	500.000	-2.5
Aroclor-1260-5	7.278	7.179	7.379	512.140	500.000	2.4
Decachlorobiphenyl	8.676	8.575	8.775	47.020	50.000	-6.0
Tetrachloro-m-xylene	3.631	3.531	3.731	50.830	50.000	1.7