



284 Sheffield Street, Mountainside, NJ 07092 Phone: 908 789 8900 Fax: 908 789 8922

CALIBRATION VERIFICATION SUMMARY

Contract: CHEM02

Lab Code: CHEM Case No.: P1187 SAS No.: P1187 SDG NO.: P1187

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 01/23/2024 01/23/2024

Client Sample No.: CCAL03 Date Analyzed: 01/29/2024

Lab Sample No.: AR1660CCC500 Data File : PO101450.D Time Analyzed: 11:48

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	5.500	5.400	5.600	501.900	500.000	0.4
Aroclor-1016-2	5.522	5.422	5.622	510.710	500.000	2.1
Aroclor-1016-3	5.584	5.485	5.685	524.480	500.000	4.9
Aroclor-1016-4	5.683	5.583	5.783	520.520	500.000	4.1
Aroclor-1016-5	5.978	5.879	6.079	482.110	500.000	-3.6
Aroclor-1260-1	7.108	7.007	7.207	461.260	500.000	-7.7
Aroclor-1260-2	7.366	7.265	7.465	482.950	500.000	-3.4
Aroclor-1260-3	7.727	7.626	7.826	470.400	500.000	-5.9
Aroclor-1260-4	7.954	7.852	8.052	491.070	500.000	-1.8
Aroclor-1260-5	8.270	8.168	8.368	480.960	500.000	-3.8
Decachlorobiphenyl	10.061	9.957	10.157	46.630	50.000	-6.7
Tetrachloro-m-xylene	4.332	4.234	4.434	47.910	50.000	-4.2



284 Sheffield Street, Mountainside, NJ 07092 Phone: 908 789 8900 Fax: 908 789 8922

CALIBRATION VERIFICATION SUMMARY

Contract: CHEM02

Lab Code: CHEM Case No.: P1187 SAS No.: P1187 SDG NO.: P1187

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 01/23/2024 01/23/2024

Client Sample No.: CCAL03 Date Analyzed: 01/29/2024

Lab Sample No.: AR1660CCC500 Data File : PO101450.D Time Analyzed: 11:48

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.722	4.623	4.823	469.100	500.000	-6.2
Aroclor-1016-2	4.741	4.642	4.842	479.960	500.000	-4.0
Aroclor-1016-3	4.920	4.821	5.021	472.830	500.000	-5.4
Aroclor-1016-4	4.960	4.862	5.062	461.790	500.000	-7.6
Aroclor-1016-5	5.176	5.077	5.277	460.920	500.000	-7.8
Aroclor-1260-1	6.217	6.117	6.317	449.190	500.000	-10.2
Aroclor-1260-2	6.406	6.306	6.506	460.440	500.000	-7.9
Aroclor-1260-3	6.561	6.461	6.661	458.170	500.000	-8.4
Aroclor-1260-4	7.035	6.935	7.135	466.940	500.000	-6.6
Aroclor-1260-5	7.278	7.179	7.379	478.530	500.000	-4.3
Decachlorobiphenyl	8.676	8.575	8.775	44.930	50.000	-10.1
Tetrachloro-m-xylene	3.630	3.531	3.731	45.800	50.000	-8.4