



284 Sheffield Street, Mountainside, NJ 07092 Phone: 908 789 8900 Fax: 908 789 8922

CALIBRATION VERIFICATION SUMMARY

Contract: CHEM02

Lab Code: CHEM Case No.: P1187 SAS No.: P1187 SDG NO.: P1187

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 01/19/2024 01/19/2024

Client Sample No.: CCAL02 Date Analyzed: 01/26/2024

Lab Sample No.: AR1660CCC500 Data File : PR065250.D Time Analyzed: 12:10

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.941	4.844	5.044	524.370	500.000	4.9
Aroclor-1016-2	4.964	4.867	5.067	534.710	500.000	6.9
Aroclor-1016-3	5.029	4.932	5.132	531.910	500.000	6.4
Aroclor-1016-4	5.133	5.035	5.235	537.820	500.000	7.6
Aroclor-1016-5	5.453	5.356	5.556	530.240	500.000	6.0
Aroclor-1260-1	6.684	6.586	6.786	536.980	500.000	7.4
Aroclor-1260-2	6.972	6.874	7.074	527.070	500.000	5.4
Aroclor-1260-3	7.367	7.269	7.469	532.350	500.000	6.5
Aroclor-1260-4	7.612	7.515	7.715	535.340	500.000	7.1
Aroclor-1260-5	7.955	7.858	8.058	532.780	500.000	6.6
Decachlorobiphenyl	9.737	9.642	9.842	53.590	50.000	7.2
Tetrachloro-m-xylene	3.701	3.603	3.803	53.130	50.000	6.3



284 Sheffield Street, Mountainside, NJ 07092 Phone: 908 789 8900 Fax: 908 789 8922

CALIBRATION VERIFICATION SUMMARY

Contract: CHEM02

Lab Code: CHEM Case No.: P1187 SAS No.: P1187 SDG NO.: P1187

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 01/19/2024 01/19/2024

Client Sample No.: CCAL02 Date Analyzed: 01/26/2024

Lab Sample No.: AR1660CCC500 Data File : PR065250.D Time Analyzed: 12:10

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.148	4.052	4.252	525.710	500.000	5.1
Aroclor-1016-2	4.167	4.070	4.270	540.840	500.000	8.2
Aroclor-1016-3	4.350	4.253	4.453	551.350	500.000	10.3
Aroclor-1016-4	4.401	4.305	4.505	547.300	500.000	9.5
Aroclor-1016-5	4.624	4.528	4.728	520.530	500.000	4.1
Aroclor-1260-1	5.739	5.643	5.843	543.880	500.000	8.8
Aroclor-1260-2	5.947	5.851	6.051	531.920	500.000	6.4
Aroclor-1260-3	6.108	6.013	6.213	534.590	500.000	6.9
Aroclor-1260-4	6.623	6.528	6.728	528.580	500.000	5.7
Aroclor-1260-5	6.891	6.797	6.997	530.800	500.000	6.2
Decachlorobiphenyl	8.373	8.281	8.481	52.650	50.000	5.3
Tetrachloro-m-xylene	3.003	2.904	3.104	53.500	50.000	7.0