



284 Sheffield Street, Mountainside, NJ 07092 Phone: 908 789 8900 Fax: 908 789 8922

INITIAL CALIBRATION OF MULTICOMPONENT ANALYTES

Contract: CHEM02

Lab Code: CHEM Case No.: P1187 SAS No.: P1187 SDG NO.: P1187

Instrument ID: ECD_L Date(s) Analyzed: 12/28/2023 12/28/2023

GC Column: ZB-MR2 ID: 0.32 (mm)

COMPOUND	AMOUNT (ng)	PEAK	RT	RT WINDOW		CALIBRATION FACTOR
				FROM	TO	
Toxaphene	500	1	6.14	6.04	6.24	29485800
		2	6.54	6.44	6.64	22559200
		3	6.97	6.87	7.07	103373000
		4	7.05	6.95	7.15	73607300
		5	7.48	7.38	7.58	57299500



284 Sheffield Street, Mountainside, NJ 07092 Phone: 908 789 8900 Fax: 908 789 8922

INITIAL CALIBRATION OF MULTICOMPONENT ANALYTES

Contract: CHEM02

Lab Code: CHEM Case No.: P1187 SAS No.: P1187 SDG NO.: P1187

Instrument ID: ECD_L Date(s) Analyzed: 12/28/2023 12/28/2023

GC Column: ZB-MR1 ID: 0.32 (mm)

COMPOUND	AMOUNT (ng)	PEAK	RT	RT WINDOW		CALIBRATION FACTOR
				FROM	TO	
Toxaphene	500	1	5.61	5.51	5.71	9147470
		2	5.89	5.79	5.99	11342700
		3	6.53	6.43	6.63	34283300
		4	6.66	6.56	6.76	47375900
		5	7.11	7.01	7.21	16492600