



NEW JERSEY LAB ID#:20012 : NEW YORK LAB ID#: 11376

GC ANALYSIS CONFORMANCE/NON-CONFORMANCE SUMMARY

CHEMTECH PROJECT NUMBER:

NA NO YES

1. Chromatograms Labeled/Compounds Identified
2. PEM Meet Criteria.
3. Calibration - Initial Calibration performed within 30 days before sample analysis and continuing calibration performed within 24 hours of sample analysis, 12 HOURS IF 8000 SERIES METHOD.
- 3(a).Initial calibration Meet Criteria
If not met, list those compounds and their recoveries which fall outside the acceptable range.
COMMENTS:

- 3(b).Continuous calibration(CCC) Meet Criteria
If not met, list those compounds and their recoveries which fall outside the acceptable range.
COMMENTS:

4. Blank Contamination - If yes, list compounds and concentrations in each blank:
COMMENTS:

5. Surrogate Recoveries Meet Criteria
If not met, list those compounds and their recoveries which fall outside the acceptable ranges.
COMMENTS:

6(a).Matrix Spike/Matrix Spike Duplicate Recoveries Meet Criteria

If not met, list those compounds and their recoveries which fall outside the acceptable range.

COMMENTS:

6(b).Blank Spike/Blank Spike Duplicate Recoveries Meet Criteria

If not met, list those compounds and their recoveries which fall outside the acceptable range.

COMMENTS:

7. Retention Time Shift Meet Criteria (if applicable)

COMMENTS:

8. Extraction Holding Time Met

If not met, list number of days exceeded for each sample:

COMMENTS:

9. Analysis Holding Time Met

If not met, list those compounds and their recoveries which fall outside the acceptable range.

COMMENTS:

ADDITIONAL COMMENTS: