



284 Sheffield Street, Mountainside, NJ 07092 Phone: 908 789 8900 Fax: 908 789 8922

### TPH GC INITIAL CALIBRATION SUMMARY

Lab Name: Chemtech Contract: CHEM02  
ProjectID: LOD-LOQ Study  
Lab Code: CHEM Case No.: P1187 SAS No.: P1187 SDG No.: P1187

| Calibration Sequence : FE021424 |            | Test : TPH GC    |                  |
|---------------------------------|------------|------------------|------------------|
| Concentration (PPM)             | Area Count | Reference Factor | File ID          |
| 1700                            | 199021138  | 117071           | FE046773.D       |
| 850                             | 101773748  | 119734           | FE046774.D       |
| 340                             | 41237873   | 121288           | FE046775.D       |
| 170                             | 22004011   | 129435           | FE046776.D       |
| 85                              | 10409611   | 122466           | FE046777.D       |
| AVG RF : 121999                 |            | % RSD : 3.788    | AVG RT : 14.8998 |