



284 Sheffield Street, Mountainside, NJ 07092 Phone: 908 789 8900 Fax: 908 789 8922

TPH GC INITIAL CALIBRATION SUMMARY

Lab Name: Chemtech Contract: CHEM02
ProjectID: LOD-LOQ Study
Lab Code: CHEM Case No.: P1187 SAS No.: P1187 SDG No.: P1187

Calibration Sequence : FG020124		Test : TPH GC	
Concentration (PPM)	Area Count	Reference Factor	File ID
1700	178623574	105073	FG013399.D
850	90750896	106766	FG013400.D
340	36346263	106901	FG013401.D
170	17859657	105057	FG013402.D
85	9195801	108186	FG013403.D
AVG RF : 106397		% RSD : 1.256	AVG RT : 15.072