



284 Sheffield Street, Mountainside, NJ 07092 Phone: 908 789 8900 Fax: 908 789 8922

RETENTION TIMES OF INITIAL CALIBRATION

Contract: CHEM02

Lab Code: CHEM Case No.: P1187 SAS No.: P1187 SDG NO.: P1187

Instrument ID: ECD_R Calibration Date(s): 01/24/2024 01/24/2024

Calibration Times: 13:28 14:38

GC Column: ZB-MR1 ID: 0.32 (mm)

LAB FILE ID:	RT 0.5 = <u>PR065164.D</u>	RT 0.25 = <u>PR065165.D</u>
RT 0.1 = <u>PR065166.D</u>	RT 0.05 = <u>PR065167.D</u>	RT 0.025 = <u>PR065168.D</u>

COMPOUND	RT 0.5	RT 0.25	RT 0.1	RT 0.05	RT 0.025	MEAN RT	RT WINDOW	
							FROM	TO
DBCP	6.69	6.69	6.69	6.69	6.69	6.69	6.59	6.79
EDB	2.90	2.90	2.90	2.89	2.90	2.90	2.80	3.00



284 Sheffield Street, Mountainside, NJ 07092 Phone: 908 789 8900 Fax: 908 789 8922

RETENTION TIMES OF INITIAL CALIBRATION

Contract: CHEM02

Lab Code: CHEM Case No.: P1187 SAS No.: P1187 SDG NO.: P1187

Instrument ID: ECD_R Calibration Date(s): 01/24/2024 01/24/2024

Calibration Times: 13:28 14:38

GC Column: ZB-MR2 ID: 0.32 (mm)

LAB FILE ID: RT 0.5 = PR065164.D RT 0.25 = PR065165.D

RT 0.1 = PR065166.D RT 0.05 = PR065167.D RT 0.025 = PR065168.D

COMPOUND	RT 0.5	RT 0.25	RT 0.1	RT 0.05	RT 0.025	MEAN RT	RT WINDOW	
							FROM	TO
DBCP	5.63	5.63	5.63	5.63	5.63	5.63	5.53	5.73
EDB	2.25	2.24	2.25	2.24	2.24	2.24	2.14	2.34