



284 Sheffield Street, Mountainside, NJ 07092 Phone: 908 789 8900 Fax: 908 789 8922

CALIBRATION VERIFICATION SUMMARY

Contract: CHEM02

Lab Code: CHEM Case No.: P1187 SAS No.: P1187 SDG NO.: P1187

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 01/24/2024 01/24/2024

Client Sample No.: CCAL01 Date Analyzed: 01/24/2024

Lab Sample No.: M8011-504 CCC 0. Data File : PR065171.D Time Analyzed: 15:08

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
DBCP	6.689	6.592	6.792	0.100	0.100	0.0
EDB	2.898	2.800	3.000	0.110	0.100	0.0



284 Sheffield Street, Mountainside, NJ 07092 Phone: 908 789 8900 Fax: 908 789 8922

CALIBRATION VERIFICATION SUMMARY

Contract: CHEM02

Lab Code: CHEM Case No.: P1187 SAS No.: P1187 SDG NO.: P1187

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 01/24/2024 01/24/2024

Client Sample No.: CCAL01 Date Analyzed: 01/24/2024

Lab Sample No.: M8011-504 CCC 0. Data File : PR065171.D Time Analyzed: 15:08

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
DBCP	5.634	5.531	5.731	0.100	0.100	0.0
EDB	2.248	2.145	2.345	0.110	0.100	0.0