



284 Sheffield Street, Mountainside, NJ 07092 Phone: 908 789 8900 Fax: 908 789 8922

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.: P1187

Client: Chemtech Consulting Group

Analytical Method: 8011

Datafile : PR065180.D

Lab Sample ID	Parameter	Spike	Result	Units	Rec	RPD	Qual	RPD	Low	Limits	RPD
								Qual		High	
PB158605BS	DBCP	0.25	0.23	ug/L	92				70	130	
	EDB	0.25	0.23	ug/L	92				70	130	