



284 Sheffield Street, Mountainside, NJ 07092 Phone: 908 789 8900 Fax: 908 789 8922

## **CASE NARRATIVE**

**LiRo Engineers, Inc.**

**Project Name: Walter Gladwin Recreation Center, Bronx, NY**

**Project # N/A**

**Chemtech Project # P1747**

**Test Name: Pesticide-TCL**

### **A. Number of Samples and Date of Receipt:**

10 Water samples were received on 03/13/2024.

### **B. Parameters**

According to the Chain of Custody document, the following analyses were requested: Anions Group1, CBOD5, Chloride, Dissolved ICP-TAL Metals, Dissolved Mercury, DISSOLVED METALS-TAL, Flash Point, Hexavalent Chromium, Mercury, Metals ICP-TAL, METALS-NYCD, METALS-TAL, Non-Polar Material, NYCD Discharge, PCB, Pesticide-TCL, Phenolics, SVOC-NYCD, SVOC-TCL BNA -20, TKN, Total Nitrogen, TS, TSS, VOC-NYCD and VOC-TCL VOA-10. This data package contains results for Pesticide-TCL.

### **C. Analytical Techniques:**

The analysis was performed on instrument ECD\_L. The front column is ZB-MR2 which is 30 meters, 0.32 mm ID, 0.25 um df, Catalog #: 7HMG017- 11 The rear column is ZB-MR1 which is 30 meters, 0.32 mm ID, 0.5 um df, Catalog # 7HM-G016-17. The analysis of Pesticide-TCLs was based on method 8081B and extraction was done based on method 3510.

### **D. QA/ QC Samples:**

The Holding Times were met for all analysis.

The Surrogate recoveries met the acceptable criteria

The Retention Times were acceptable for all samples.

The RPD met criteria .

The Blank Spike met requirements for all samples .

The Blank Spike Duplicate met requirements for all samples .

The Blank analysis did not indicate the presence of lab contamination.

The Initial Calibration met the requirements .

The Continuous Calibration met the requirements .

### **E. Additional Comments:**

### **F. Manual Integration Comments:**

Please refer to the Manual integration Report included with the Run Logs for information on the manual integrations performed.

---



284 Sheffield Street, Mountainside, NJ 07092 Phone: 908 789 8900 Fax: 908 789 8922

I certify that the data package is in compliance with the terms and conditions of the contract, both technically and for completeness, for other than the conditions detailed above. The laboratory manager or his designee, as verified by the following signature has authorized release of the data contained in this hard copy data package.

Signature\_\_\_\_\_