

Analytical Summary Report

Analysis Method: 1010B

Reviewed By: Nikita

Parameter: Flash Point

Supervisor Review By: Iwona

Run Number: LB129879

Ambient Barometric Pressure (mmHg): 771.00

Thermometer ID: Flashpoint

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W3080

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	8	28.00	82.4	81.7	03/14/2024	12:45
2	P1747-03		1	13	100.00	>212.0	>212.0	03/14/2024	13:15
3	P1747-03DUP		1	12	100.00	>212.0	>212.0	03/14/2024	13:45

$$\text{Result} = (\text{Celsius} * 1.8) + 32$$
$$\text{Final Result} = \text{Result} + (760 - \text{Ambient Barometric Pressure}) * 0.06$$

LB. 129879

WORKLIST(Hardcopy Internal Chain)

WorkList Name : 03/14/24 flashpoint

WorkList ID : 178584

Department : Wet-Chemistry

Date : 03-14-2024 11:12:29

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
P1747-03	MW-01	Water	Flash Point	Cool 4 deg C	LIRO01	I31	03/13/2024	1010B

Date/Time 03/14/2024 12:30
Raw Sample Received by: N.P.W.C
Raw Sample Relinquished by: JH(CDC)

Date/Time 03/14/2024 14:00
Raw Sample Received by: JH(CDC)
Raw Sample Relinquished by: N.P.W.C