



284 Sheffield Street, Mountainside, NJ 07092 Phone: 908 789 8900 Fax: 908 789 8922

Instrument ID: P4

Daily Analysis Runlog For Sequence/QC Batch ID # LB129894

Review By	jaswal	Review On	3/19/2024 2:53:55 AM
Supervise By	mohan	Supervise On	3/19/2024 3:35:54 AM

STD. NAME	STD REF.#
ICAL Standard	MP79389,MP79360,MP79397,MP79396,MP79395,MP79394
ICV Standard	MP79393
CCV Standard	MP79398
ICSA Standard	MP79404,MP79405
CRI Standard	MP79360
LCS Standard	
Chk Standard	MP79404,MP79402

Sr#	SampleID	ClientID	QcType	Date	Comment	Operator	Status
1	S0	S0	CAL1	03/14/24 12:06		Bin	OK
2	S1	S1	CAL2	03/14/24 12:11		Bin	OK
3	S2	S2	CAL3	03/14/24 12:15		Bin	OK
4	S3	S3	CAL4	03/14/24 12:19		Bin	OK
5	S4	S4	CAL5	03/14/24 12:23		Bin	OK
6	S5	S5	CAL6	03/14/24 12:28		Bin	OK
7	ICV01	ICV01	ICV	03/14/24 12:40		Bin	OK
8	LLICV01	LLICV01	LLICV	03/14/24 12:46		Bin	OK
9	ICB01	ICB01	ICB	03/14/24 12:51		Bin	OK
10	CRI01	CRI01	CRDL	03/14/24 12:56		Bin	OK
11	ICSA01	ICSA01	ICSA	03/14/24 13:00		Bin	OK
12	ICSAB01	ICSAB01	ICSAB	03/14/24 13:04		Bin	OK
13	ICSADL	ICSADL	ICSA	03/14/24 13:09		Bin	OK
14	ICSABDL	ICSABDL	ICSAB	03/14/24 13:13		Bin	OK
15	CCV01	CCV01	CCV	03/14/24 13:18		Bin	OK
16	CCB01	CCB01	CCB	03/14/24 13:22		Bin	OK
17	PB159570BL	PB159570BL	MB	03/14/24 13:31		Bin	OK
18	PB159570BS	PB159570BS	LCS	03/14/24 13:35		Bin	OK
19	P1726-01	M6-CONCRETE	SAM	03/14/24 13:40		Bin	OK
20	P1733-01	WC-1	SAM	03/14/24 13:44		Bin	OK



284 Sheffield Street, Mountainside, NJ 07092 Phone: 908 789 8900 Fax: 908 789 8922

Instrument ID: P4

Daily Analysis Runlog For Sequence/QC Batch ID # LB129894

Review By	jaswal	Review On	3/19/2024 2:53:55 AM
Supervise By	mohan	Supervise On	3/19/2024 3:35:54 AM

STD. NAME	STD REF.#
ICAL Standard	MP79389,MP79360,MP79397,MP79396,MP79395,MP79394
ICV Standard	MP79393
CCV Standard	MP79398
ICSA Standard	MP79404,MP79405
CRI Standard	MP79360
LCS Standard	
Chk Standard	MP79404,MP79402

21	P1733-05	WC-2	SAM	03/14/24 13:48		Bin	OK
22	P1733-05DUP	WC-2DUP	DUP	03/14/24 13:53		Bin	OK
23	P1733-05L	WC-2L	SD	03/14/24 13:57		Bin	OK
24	P1733-05MS	WC-2MS	MS	03/14/24 14:01		Bin	OK
25	P1733-05MSD	WC-2MSD	MSD	03/14/24 14:05		Bin	OK
26	P1733-05A	WC-2A	PS	03/14/24 14:09		Bin	OK
27	CCV02	CCV02	CCV	03/14/24 14:14		Bin	OK
28	CCB02	CCB02	CCB	03/14/24 14:18		Bin	OK
29	P1733-09	WC-3	SAM	03/14/24 14:22		Bin	OK
30	P1744-01	RT-3488	SAM	03/14/24 14:27		Bin	OK
31	PB159541BL	PB159541BL	MB	03/14/24 14:31		Bin	OK
32	PB159541BS	PB159541BS	LCS	03/14/24 14:35		Bin	OK
33	P1714-04	WATER-TREATMENT	SAM	03/14/24 14:47	Not Use	Bin	Not Ok
34	P1714-04DUP	WATER-TREATMENT	DUP	03/14/24 14:51	Not Use	Bin	Not Ok
35	P1714-04L	WATER-TREATMENT	SD	03/14/24 14:56	Not Use	Bin	Not Ok
36	P1714-04MS	WATER-TREATMENT	MS	03/14/24 15:00	Not Use	Bin	Not Ok
37	P1714-04MSD	WATER-TREATMENT	MSD	03/14/24 15:05	Not Use	Bin	Not Ok
38	P1714-04A	WATER-TREATMENT	PS	03/14/24 15:09	Not Use	Bin	Not Ok
39	CCV03	CCV03	CCV	03/14/24 15:27		Bin	OK
40	CCB03	CCB03	CCB	03/14/24 15:32		Bin	OK
41	P1734-06	COMP-2	SAM	03/14/24 15:50		Bin	OK



Instrument ID: P4

Daily Analysis Runlog For Sequence/QC Batch ID # LB129894

Review By		jaswal	Review On		3/19/2024 2:53:55 AM		
Supervise By		mohan	Supervise On		3/19/2024 3:35:54 AM		
STD. NAME		STD REF.#					
ICAL Standard		MP79389,MP79360,MP79397,MP79396,MP79395,MP79394					
ICV Standard		MP79393					
CCV Standard		MP79398					
ICSA Standard		MP79404,MP79405					
CRI Standard		MP79360					
LCS Standard							
Chk Standard		MP79404,MP79402					
42	P1725-01	RT-3860	SAM	03/14/24 15:55		Bin	OK
43	LR1	LR1	HIGH STD	03/14/24 15:59		Bin	OK
44	LR2	LR2	HIGH STD	03/14/24 16:10		Bin	OK
45	CCV04	CCV04	CCV	03/14/24 16:45		Bin	OK
46	CCB04	CCB04	CCB	03/14/24 16:56		Bin	OK



284 Sheffield Street, Mountainside, NJ 07092 Phone: 908 789 8900 Fax: 908 789 8922

Instrument ID: P4

Daily Analysis Runlog For Sequence/QC Batch ID # LB129934

Review By	mohan	Review On	3/27/2024 4:28:31 PM
Supervise By	Kareem	Supervise On	3/27/2024 4:32:26 PM

STD. NAME	STD REF.#
ICAL Standard	MP79389,MP79360,MP79397,MP79396,MP79395,MP79394
ICV Standard	MP79393
CCV Standard	MP79398
ICSA Standard	MP79404,MP79405
CRI Standard	MP79360
LCS Standard	
Chk Standard	MP79404,MP79402

Sr#	SampleId	ClientID	QcType	Date	Comment	Operator	Status
1	S0	S0	CAL1	03/18/24 16:53		Bin	OK
2	S1	S1	CAL2	03/18/24 16:58		Bin	OK
3	S2	S2	CAL3	03/18/24 17:02		Bin	OK
4	S3	S3	CAL4	03/18/24 17:06		Bin	OK
5	S4	S4	CAL5	03/18/24 17:11		Bin	OK
6	S5	S5	CAL6	03/18/24 17:15		Bin	OK
7	ICV01	ICV01	ICV	03/18/24 17:48	ICV Fail for Co,Fe,Ni (95-105) (200.7)	Bin	OK
8	LLICV01	LLICV01	LLICV	03/18/24 17:54		Bin	OK
9	ICB01	ICB01	ICB	03/18/24 17:58		Bin	OK
10	CRI01	CRI01	CRDL	03/18/24 18:03		Bin	OK
11	ICSA01	ICSA01	ICSA	03/18/24 18:07		Bin	OK
12	ICSAB01	ICSAB01	ICSAB	03/18/24 18:21		Bin	OK
13	ICSADL	ICSADL	ICSA	03/18/24 18:25		Bin	OK
14	ICSABDL	ICSABDL	ICSAB	03/18/24 18:29		Bin	OK
15	CCV01	CCV01	CCV	03/18/24 18:34		Bin	OK
16	CCB01	CCB01	CCB	03/18/24 18:38		Bin	OK
17	PB159517BL	PB159517BL	MB	03/18/24 18:51		Bin	OK
18	PB159517BS	PB159517BS	LCS	03/18/24 18:55		Bin	OK
19	P1675-02	SPRING-2024	SAM	03/18/24 18:59	ICV Fail For Ni	Bin	Not Ok



Instrument ID: P4

Daily Analysis Runlog For Sequence/QC Batch ID # LB129934

Review By	mohan	Review On	3/27/2024 4:28:31 PM
Supervise By	Kareem	Supervise On	3/27/2024 4:32:26 PM

STD. NAME	STD REF.#
ICAL Standard	MP79389,MP79360,MP79397,MP79396,MP79395,MP79394
ICV Standard	MP79393
CCV Standard	MP79398
ICSA Standard	MP79404,MP79405
CRI Standard	MP79360
LCS Standard	
Chk Standard	MP79404,MP79402

20	P1723-02	EFF-WW	SAM	03/18/24 19:04	ICV Fail For Ni	Bin	Not Ok
21	P1747-03DUP	MW-01DUP	DUP	03/18/24 19:21	ICV Fail For Ni	Bin	Not Ok
22	P1747-03	MW-01	SAM	03/18/24 19:26	ICV Fail For Ni	Bin	Not Ok
23	P1747-03L	MW-01L	SD	03/18/24 19:30	ICV Fail For Ni	Bin	Not Ok
24	P1747-03MS	MW-01MS	MS	03/18/24 19:35	ICV Fail For Ni	Bin	Not Ok
25	P1747-03MSD	MW-01MSD	MSD	03/18/24 19:39	ICV Fail For Ni	Bin	Not Ok
26	P1747-03A	MW-01A	PS	03/18/24 19:43	ICV Fail For Ni	Bin	Not Ok
27	CCV02	CCV02	CCV	03/18/24 19:48		Bin	OK
28	CCB02	CCB02	CCB	03/18/24 19:52		Bin	OK
29	PB159617BL	PB159617BL	MB	03/18/24 19:56		Bin	OK
30	PB159617BS	PB159617BS	LCS	03/18/24 20:01		Bin	OK
31	P1764-10	COMP-1	SAM	03/18/24 20:05		Bin	OK
32	P1764-10DUP	COMP-1DUP	DUP	03/18/24 20:09		Bin	OK
33	P1764-10L	COMP-1L	SD	03/18/24 20:13		Bin	OK
34	P1764-10MS	COMP-1MS	MS	03/18/24 20:18		Bin	OK
35	P1764-10MSD	COMP-1MSD	MSD	03/18/24 20:22		Bin	OK
36	P1764-10A	COMP-1A	PS	03/18/24 20:26		Bin	OK
37	P1764-11	COMP-2	SAM	03/18/24 20:30		Bin	OK
38	P1767-01	NB-07-031424	SAM	03/18/24 20:34		Bin	OK
39	CCV03	CCV03	CCV	03/18/24 21:09		Bin	OK
40	CCB03	CCB03	CCB	03/18/24 21:13		Bin	OK



284 Sheffield Street, Mountainside, NJ 07092 Phone: 908 789 8900 Fax: 908 789 8922

Instrument ID: P4

Daily Analysis Runlog For Sequence/QC Batch ID # LB129934

Review By	mohan	Review On	3/27/2024 4:28:31 PM
Supervise By	Kareem	Supervise On	3/27/2024 4:32:26 PM

STD. NAME	STD REF.#
ICAL Standard	MP79389,MP79360,MP79397,MP79396,MP79395,MP79394
ICV Standard	MP79393
CCV Standard	MP79398
ICSA Standard	MP79404,MP79405
CRI Standard	MP79360
LCS Standard	
Chk Standard	MP79404,MP79402

41	P1772-01	1151	SAM	03/18/24 21:17		Bin	OK
42	P1772-02	1276	SAM	03/18/24 21:22		Bin	OK
43	P1774-01	TP3	SAM	03/18/24 21:26		Bin	OK
44	P1774-02	TP9	SAM	03/18/24 21:30		Bin	OK
45	LR1	LR1	HIGH STD	03/18/24 21:35		Bin	OK
46	LR2	LR2	HIGH STD	03/18/24 21:45	Need To Confirm For Bed Injection	Bin	OK
47	P1747-01	MW-01	SAM	03/18/24 21:51		Bin	OK
48	P1747-02	MW-01-DUP	SAM	03/18/24 21:55		Bin	OK
49	P1747-04	MW-02	SAM	03/18/24 22:00		Bin	OK
50	P1747-05	TWP-04	SAM	03/18/24 22:04		Bin	OK
51	CCV04	CCV04	CCV	03/18/24 22:27		Bin	OK
52	CCB04	CCB04	CCB	03/18/24 22:31		Bin	OK
53	P1747-07	MW-01	SAM	03/18/24 22:36	Need to Confirm	Bin	OK
54	P1747-08	MW-01-DUP	SAM	03/18/24 22:40		Bin	OK
55	P1747-09	MW-02	SAM	03/18/24 22:44	Need to Confirm	Bin	OK
56	P1747-10	TWP-04	SAM	03/18/24 22:49	Need to Confirm	Bin	OK
57	P1749-01	MW-1	SAM	03/18/24 22:53		Bin	OK
58	P1749-02	MW-2A	SAM	03/18/24 22:58		Bin	OK
59	P1749-03	MW-2	SAM	03/18/24 23:02		Bin	OK
60	P1749-04	MW-4	SAM	03/18/24 23:07		Bin	OK
61	PB159591BL	PB159591BL	MB	03/18/24 23:11		Bin	OK



Instrument ID: P4

Daily Analysis Runlog For Sequence/QC Batch ID # LB129934

Review By	mohan	Review On	3/27/2024 4:28:31 PM
Supervise By	Kareem	Supervise On	3/27/2024 4:32:26 PM

STD. NAME	STD REF.#
ICAL Standard	MP79389,MP79360,MP79397,MP79396,MP79395,MP79394
ICV Standard	MP79393
CCV Standard	MP79398
ICSA Standard	MP79404,MP79405
CRI Standard	MP79360
LCS Standard	
Chk Standard	MP79404,MP79402

62	PB159591BS	PB159591BS	LCS	03/18/24 23:16		Bin	OK
63	CCV05	CCV05	CCV	03/18/24 23:20		Bin	OK
64	CCB05	CCB05	CCB	03/18/24 23:24		Bin	OK
65	P1746-01	SB-01-0-2.0	SAM	03/18/24 23:42		Bin	OK
66	P1746-02	SB-01-4.0-6.0	SAM	03/18/24 23:47		Bin	OK
67	P1746-03	SB-01-4.0-6.0-DUP	SAM	03/18/24 23:51		Bin	OK
68	P1746-04RE	SB-01-7.0-9.5RE	SAM	03/18/24 23:56	Not Required	Bin	Not Ok
69	P1746-06	SB-02-0-2.0	SAM	03/19/24		Bin	OK
70	P1746-08	SB-03-0-2.0	SAM	03/19/24 00:05		Bin	OK
71	P1746-09	SB-03-2.0-3.0	SAM	03/19/24 00:09		Bin	OK
72	P1746-11	SB-04-0-2.0	SAM	03/19/24 00:14		Bin	OK
73	P1746-12	SB-04-2.0-4.0	SAM	03/19/24 00:18		Bin	OK
74	P1746-13	SB-04-4.0-6.0	SAM	03/19/24 00:22		Bin	OK
75	CCV06	CCV06	CCV	03/19/24 00:42		Bin	OK
76	CCB06	CCB06	CCB	03/19/24 00:51		Bin	OK
77	P1746-15	SB-05-0-2.0	SAM	03/19/24 00:56		Bin	OK
78	P1746-16	SB-05-0-2.0-DUP	SAM	03/19/24 01:00		Bin	OK
79	P1746-17	SB-05-2.0-4.5	SAM	03/19/24 01:04		Bin	OK
80	P1746-19	SB-06-0-2.0	SAM	03/19/24 01:09		Bin	OK
81	P1746-19DUP	SB-06-0-2.0DUP	DUP	03/19/24 01:13		Bin	OK
82	P1746-19L	SB-06-0-2.0L	SD	03/19/24 01:17		Bin	OK



284 Sheffield Street, Mountainside, NJ 07092 Phone: 908 789 8900 Fax: 908 789 8922

Instrument ID: P4

Daily Analysis Runlog For Sequence/QC Batch ID # LB129934

Review By	mohan	Review On	3/27/2024 4:28:31 PM
Supervise By	Kareem	Supervise On	3/27/2024 4:32:26 PM

STD. NAME	STD REF.#
ICAL Standard	MP79389,MP79360,MP79397,MP79396,MP79395,MP79394
ICV Standard	MP79393
CCV Standard	MP79398
ICSA Standard	MP79404,MP79405
CRI Standard	MP79360
LCS Standard	
Chk Standard	MP79404,MP79402

83	P1746-19MS	SB-06-0-2.0MS	MS	03/19/24 01:36		Bin	OK
84	P1746-19MSD	SB-06-0-2.0MSD	MSD	03/19/24 01:40		Bin	OK
85	P1746-19A	SB-06-0-2.0A	PS	03/19/24 01:44		Bin	OK
86	P1746-04	SB-01-7.0-9.5	SAM	03/19/24 01:48		Bin	OK
87	CCV07	CCV07	CCV	03/19/24 01:53		Bin	OK
88	CCB07	CCB07	CCB	03/19/24 01:57		Bin	OK



284 Sheffield Street, Mountainside, NJ 07092 Phone: 908 789 8900 Fax: 908 789 8922

Instrument ID: CV1

Daily Analysis Runlog For Sequence/QC Batch ID # LB129940

Review By	jaswal	Review On	3/19/2024 10:54:26 PM
Supervise By	mohan	Supervise On	3/20/2024 3:30:16 AM

STD. NAME	STD REF.#
ICAL Standard	MP79805,MP79806,MP79807,MP79808,MP79809,MP79810
ICV Standard	MP79811
CCV Standard	MP79813
ICSA Standard	
CRI Standard	MP79816
LCS Standard	
Chk Standard	MP79812,MP79814,MP79817,MP79836

Sr#	SampleId	ClientID	QcType	Date	Comment	Operator	Status
1	S0	S0	CAL1	03/19/24 12:17		mohan	OK
2	S0.2	S0.2	CAL2	03/19/24 12:20		mohan	OK
3	S2.5	S2.5	CAL3	03/19/24 12:22		mohan	OK
4	S5	S5	CAL4	03/19/24 12:24		mohan	OK
5	S7.5	S7.5	CAL5	03/19/24 12:26		mohan	OK
6	S10	S10	CAL6	03/19/24 12:31		mohan	OK
7	ICV53	ICV53	ICV	03/19/24 12:37		mohan	OK
8	ICB53	ICB53	ICB	03/19/24 12:39		mohan	OK
9	CCV81	CCV81	CCV	03/19/24 12:42		mohan	OK
10	CCB81	CCB81	CCB	03/19/24 12:44		mohan	OK
11	CRA	CRA	CRDL	03/19/24 12:46		mohan	OK
12	HighStd	HighStd	HIGH STD	03/19/24 12:48		mohan	OK
13	ChkStd	ChkStd	SAM	03/19/24 12:51		mohan	OK
14	PB159624BL	PB159624BL	MB	03/19/24 12:56		mohan	OK
15	PB159624BS	PB159624BS	LCS	03/19/24 12:58		mohan	OK
16	P1746-05	SB-01-COMP	SAM	03/19/24 13:00		mohan	OK
17	P1746-07	SB-02-COMP	SAM	03/19/24 13:02		mohan	OK
18	P1746-10	SB-03-COMP	SAM	03/19/24 13:05		mohan	OK
19	P1746-14	SB-04-COMP	SAM	03/19/24 13:07		mohan	OK
20	P1746-18	SB-05-COMP	SAM	03/19/24 13:09		mohan	OK



284 Sheffield Street, Mountainside, NJ 07092 Phone: 908 789 8900 Fax: 908 789 8922

Instrument ID: CV1

Daily Analysis Runlog For Sequence/QC Batch ID # LB129940

Review By	jaswal	Review On	3/19/2024 10:54:26 PM
Supervise By	mohan	Supervise On	3/20/2024 3:30:16 AM

STD. NAME	STD REF.#
ICAL Standard	MP79805,MP79806,MP79807,MP79808,MP79809,MP79810
ICV Standard	MP79811
CCV Standard	MP79813
ICSA Standard	
CRI Standard	MP79816
LCS Standard	
Chk Standard	MP79812,MP79814,MP79817,MP79836

21	CCV82	CCV82	CCV	03/19/24 13:12		mohan	OK
22	CCB82	CCB82	CCB	03/19/24 13:14		mohan	OK
23	P1746-20	SB-06-COMP	SAM	03/19/24 13:16		mohan	OK
24	P1771-02	COMP-1	SAM	03/19/24 13:18		mohan	OK
25	P1771-04	COMP-2	SAM	03/19/24 13:21		mohan	OK
26	P1771-06	COMP-3	SAM	03/19/24 13:23		mohan	OK
27	P1771-08	COMP-4	SAM	03/19/24 13:25		mohan	OK
28	P1771-08DUP	COMP-4DUP	DUP	03/19/24 13:27		mohan	OK
29	P1771-08MS	COMP-4MS	MS	03/19/24 13:32		mohan	OK
30	P1771-08MSD	COMP-4MSD	MSD	03/19/24 13:35		mohan	OK
31	PB159659BL	PB159659BL	MB	03/19/24 13:37		mohan	OK
32	PB159659BS	PB159659BS	LCS	03/19/24 13:39		mohan	OK
33	CCV83	CCV83	CCV	03/19/24 13:42		mohan	OK
34	CCB83	CCB83	CCB	03/19/24 13:44		mohan	OK
35	P1747-01	MW-01	SAM	03/19/24 13:46		mohan	OK
36	P1747-02	MW-01-DUP	SAM	03/19/24 13:48		mohan	OK
37	P1747-04	MW-02	SAM	03/19/24 13:51		mohan	OK
38	P1747-05	TWP-04	SAM	03/19/24 13:53		mohan	OK
39	P1747-07	MW-01	SAM	03/19/24 13:55		mohan	OK
40	P1747-08	MW-01-DUP	SAM	03/19/24 13:58		mohan	OK
41	P1747-09	MW-02	SAM	03/19/24 14:00		mohan	OK



284 Sheffield Street, Mountainside, NJ 07092 Phone: 908 789 8900 Fax: 908 789 8922

Instrument ID: CV1

Daily Analysis Runlog For Sequence/QC Batch ID # LB129940

Review By	jaswal	Review On	3/19/2024 10:54:26 PM
Supervise By	mohan	Supervise On	3/20/2024 3:30:16 AM

STD. NAME	STD REF.#
ICAL Standard	MP79805,MP79806,MP79807,MP79808,MP79809,MP79810
ICV Standard	MP79811
CCV Standard	MP79813
ICSA Standard	
CRI Standard	MP79816
LCS Standard	
Chk Standard	MP79812,MP79814,MP79817,MP79836

42	P1747-10	TWP-04	SAM	03/19/24 14:02		mohan	OK
43	P1769-01	MW-2	SAM	03/19/24 14:04		mohan	OK
44	P1769-02	MW-1	SAM	03/19/24 14:07		mohan	OK
45	CCV84	CCV84	CCV	03/19/24 14:09		mohan	OK
46	CCB84	CCB84	CCB	03/19/24 14:11		mohan	OK
47	P1769-03	FB-3-14-24	SAM	03/19/24 14:13		mohan	OK
48	P1772-06	1350	SAM	03/19/24 14:16		mohan	OK
49	P1772-08	1346	SAM	03/19/24 14:23		mohan	OK
50	P1772-08DUP	1346DUP	DUP	03/19/24 14:27		mohan	OK
51	P1772-08MS	1346MS	MS	03/19/24 14:32		mohan	OK
52	P1772-08MSD	1346MSD	MSD	03/19/24 14:35		mohan	OK
53	CCV85	CCV85	CCV	03/19/24 14:40		mohan	OK
54	CCB85	CCB85	CCB	03/19/24 14:42		mohan	OK
55	PB159589TB	PB159589TB	MB	03/19/24 14:44		mohan	OK
56	P1771-08L	COMP-4L	SD	03/19/24 14:47		mohan	OK
57	P1771-08A	COMP-4A	PS	03/19/24 14:49		mohan	OK
58	P1772-08L	1346L	SD	03/19/24 14:52		mohan	OK
59	P1772-08A	1346A	PS	03/19/24 14:54		mohan	OK
60	CCV86	CCV86	CCV	03/19/24 14:57		mohan	OK
61	CCB86	CCB86	CCB	03/19/24 14:59		mohan	OK



284 Sheffield Street, Mountainside, NJ 07092 Phone: 908 789 8900 Fax: 908 789 8922

Instrument ID: CV1

Daily Analysis Runlog For Sequence/QC Batch ID # LB129958

Review By	jaswal	Review On	3/20/2024 10:21:07 PM
Supervise By		Supervise On	

STD. NAME	STD REF.#
ICAL Standard	MP79853,MP79854,MP79855,MP79856,MP79857,MP79858,MP79859
ICV Standard	MP79860
CCV Standard	MP79862
ICSA Standard	
CRI Standard	MP79816
LCS Standard	
Chk Standard	MP79861,MP79863,MP79865,MP79850

Sr#	SampleId	ClientID	QcType	Date	Comment	Operator	Status
1	S0	S0	CAL1	03/20/24 13:55		mohan	OK
2	S0.05	S0.05	CAL2	03/20/24 13:58		mohan	OK
3	S0.2	S0.2	CAL3	03/20/24 14:00		mohan	OK
4	S2.5	S2.5	CAL4	03/20/24 14:04		mohan	OK
5	S5	S5	CAL5	03/20/24 14:06		mohan	OK
6	S7.5	S7.5	CAL6	03/20/24 14:09		mohan	OK
7	S10	S10	CAL7	03/20/24 14:16		mohan	OK
8	ICV55	ICV55	ICV	03/20/24 14:24		mohan	OK
9	ICB55	ICB55	ICB	03/20/24 14:26		mohan	OK
10	CCV89	CCV89	CCV	03/20/24 14:33		mohan	OK
11	CCB89	CCB89	CCB	03/20/24 14:35		mohan	OK
12	CRA	CRA	CRDL	03/20/24 14:39		mohan	OK
13	HighStd	HighStd	HIGH STD	03/20/24 14:42		mohan	OK
14	ChkStd	ChkStd	SAM	03/20/24 14:44		mohan	OK
15	PB159691BL	PB159691BL	MB	03/20/24 14:52		mohan	OK
16	PB159691BS	PB159691BS	LCS	03/20/24 14:59		mohan	OK
17	P1187-10	LOD-MDL-WATER-04	SAM	03/20/24 15:10		mohan	OK
18	P1187-10RE	LOD-MDL-WATER-04	SAM	03/20/24 15:12		mohan	OK
19	P1187-12	MDL-WATER-06-QT1	SAM	03/20/24 15:16		mohan	OK
20	P1747-03	MW-01	SAM	03/20/24 15:19		mohan	OK



284 Sheffield Street, Mountainside, NJ 07092 Phone: 908 789 8900 Fax: 908 789 8922

Instrument ID: CV1

Daily Analysis Runlog For Sequence/QC Batch ID # LB129958

Review By	jaswal	Review On	3/20/2024 10:21:07 PM
Supervise By		Supervise On	

STD. NAME	STD REF.#
ICAL Standard	MP79853,MP79854,MP79855,MP79856,MP79857,MP79858,MP79859
ICV Standard	MP79860
CCV Standard	MP79862
ICSA Standard	
CRI Standard	MP79816
LCS Standard	
Chk Standard	MP79861,MP79863,MP79865,MP79850

21	P1788-01	PEZOMETER-200	SAM	03/20/24 15:22		mohan	OK
22	CCV90	CCV90	CCV	03/20/24 15:24		mohan	OK
23	CCB90	CCB90	CCB	03/20/24 15:26		mohan	OK
24	P1788-02	PEZOMETER-206	SAM	03/20/24 15:29		mohan	OK
25	P1788-02DUP	PEZOMETER-206DU	DUP	03/20/24 15:31		mohan	OK
26	P1788-02MS	PEZOMETER-206MS	MS	03/20/24 15:35		mohan	OK
27	P1788-02MSD	PEZOMETER-206MS	MSD	03/20/24 15:37		mohan	OK
28	P1788-02L	PEZOMETER-206L	SD	03/20/24 15:40		mohan	OK
29	P1788-02A	PEZOMETER-206A	PS	03/20/24 15:43		mohan	OK
30	CCV91	CCV91	CCV	03/20/24 15:47		mohan	OK
31	CCB91	CCB91	CCB	03/20/24 15:50		mohan	OK



284 Sheffield Street, Mountainside, NJ 07092 Phone: 908 789 8900 Fax: 908 789 8922

Instrument ID: P4

Daily Analysis Runlog For Sequence/QC Batch ID # LB129987

Review By		Review On					
Supervise By		Supervise On					
STD. NAME		STD REF.#					
ICAL Standard ICV Standard CCV Standard ICSA Standard CRI Standard LCS Standard Chk Standard							

Sr#	SampleID	ClientID	QcType	Date	Comment	Operator	Status
1	S0	S0	CAL1	03/21/24 10:56		Bin	OK
2	S1	S1	CAL2	03/21/24 11:01		Bin	OK
3	S2	S2	CAL3	03/21/24 11:05		Bin	OK
4	S3	S3	CAL4	03/21/24 11:09		Bin	OK
5	S4	S4	CAL5	03/21/24 11:15		Bin	OK
6	S5	S5	CAL6	03/21/24 11:19		Bin	OK
7	ICV01	ICV01	ICV	03/21/24 11:59	ICV Fail for Be,Co,Fe,Mn,S,Tl (95-105) (200.7) & S (90 - 110) (6010D)	Bin	OK
8	LLICV01	LLICV01	LLICV	03/21/24 12:03	LLICV Fail For Cu,S	Bin	OK
9	ICB01	ICB01	ICB	03/21/24 12:07		Bin	OK
10	CRI01	CRI01	CRDL	03/21/24 12:12		Bin	OK
11	ICSA01	ICSA01	ICSA	03/21/24 12:16		Bin	OK
12	ICSAB01	ICSAB01	ICSAB	03/21/24 12:21		Bin	OK
13	ICSADL	ICSADL	ICSA	03/21/24 12:25		Bin	OK
14	ICSABDL	ICSABDL	ICSAB	03/21/24 12:29		Bin	OK
15	CCV01	CCV01	CCV	03/21/24 12:34		Bin	OK
16	CCB01	CCB01	CCB	03/21/24 12:38		Bin	OK
17	P1675-02	SPRING-2024	SAM	03/21/24 12:43		Bin	OK
18	P1792-01	72-11977	SAM	03/21/24 12:47		Bin	OK

Instrument ID: P4

Daily Analysis Runlog For Sequence/QC Batch ID # LB129987

Review By		Review On					
Supervise By		Supervise On					
STD. NAME	STD REF.#						
ICAL Standard ICV Standard CCV Standard ICSA Standard CRI Standard LCS Standard Chk Standard							

19	P1793-02	MS	MS	03/21/24 12:52		Bin	OK
20	P1792-01DL	72-11977DL	SAM	03/21/24 13:03	Not Required	Bin	Not Ok
21	P1792-01DL2	72-11977DL2	SAM	03/21/24 13:08	Not Required	Bin	Not Ok
22	P1772-06DL	1350DL	SAM	03/21/24 13:12		Bin	OK
23	P1772-08DL	1346DL	SAM	03/21/24 13:16		Bin	OK
24	P1772-06DL2	1350DL2	SAM	03/21/24 13:21	Not Required	Bin	Not Ok
25	P1772-06DL3	1350DL	SAM	03/21/24 13:25	Not Required	Bin	Not Ok
26	PB159589TB	PB159589TB	MB	03/21/24 13:30		Bin	OK
27	CCV02	CCV02	CCV	03/21/24 13:34		Bin	OK
28	CCB02	CCB02	CCB	03/21/24 13:39		Bin	OK
29	P1747-03	MW-01	SAM	03/21/24 13:43		Bin	OK
30	P1747-03DUP	MW-01DUP	DUP	03/21/24 13:48		Bin	OK
31	P1747-03L	MW-01L	SD	03/21/24 13:52		Bin	OK
32	P1747-03MS	MW-01MS	MS	03/21/24 13:56		Bin	OK
33	P1747-03MSD	MW-01MSD	MSD	03/21/24 14:00		Bin	OK
34	P1747-03A	MW-01A	PS	03/21/24 14:04		Bin	OK
35	P1714-04	WATER-TREATMENT	SAM	03/21/24 14:09		Bin	OK
36	P1714-04DUP	WATER-TREATMENT	DUP	03/21/24 14:13	K,Na High	Bin	Dilution
37	P1714-04L	WATER-TREATMENT	SD	03/21/24 14:18	Na High	Bin	Dilution
38	CCV03	CCV03	CCV	03/21/24 14:27		Bin	OK
39	CCB03	CCB03	CCB	03/21/24 14:31		Bin	OK



Instrument ID: P4

Daily Analysis Runlog For Sequence/QC Batch ID # LB129987

Review By		Review On					
Supervise By		Supervise On					
STD. NAME		STD REF.#					
ICAL Standard ICV Standard CCV Standard ICSA Standard CRI Standard LCS Standard Chk Standard							
40	P1714-04MSD	WATER-TREATMENT	MSD	03/21/24 14:39	K,Na High	Bin	Dilution
41	P1714-04A	WATER-TREATMENT	PS	03/21/24 14:43		Bin	OK
42	PB159618BL	PB159618BL	MB	03/21/24 14:47		Bin	OK
43	PB159618BS	PB159618BS	LCS	03/21/24 14:54		Bin	OK
44	P1786-01	TP-15	SAM	03/21/24 14:58	ICV Fail For S	Bin	Not Ok
45	P1786-01DUP	TP-15DUP	DUP	03/21/24 15:02	ICV Fail For S	Bin	Not Ok
46	P1786-01L	TP-15L	SD	03/21/24 15:06	ICV Fail For S	Bin	Not Ok
47	CCV04	CCV04	CCV	03/21/24 15:23		Bin	OK
48	CCB04	CCB04	CCB	03/21/24 15:28		Bin	OK
49	P1791-01	SU-04-03182024	SAM	03/21/24 15:32		Bin	OK
50	P1801-01	OR-03-031924	SAM	03/21/24 15:37		Bin	OK
51	P1803-01	TRENCH-PIT	SAM	03/21/24 15:41		Bin	OK
52	P1804-01	JC-03-03192024-A	SAM	03/21/24 15:45		Bin	OK
53	P1804-03	JC-03-03192024-B	SAM	03/21/24 15:49		Bin	OK
54	P1714-04MS	WATER-TREATMENT	MS	03/21/24 15:53	K,Na High	Bin	Dilution
55	PB159641BL	PB159641BL	MB	03/21/24 15:58	CCV Fail For S	Bin	Not Ok
56	P1800-01	1394	SAM	03/21/24 16:07		Bin	OK
57	PB159641BS	PB159641BS	LCS	03/21/24 16:15	CCV Fail For S	Bin	Not Ok
58	P1816-01	PL-01-032024	SAM	03/21/24 16:19		Bin	OK
59	CCV05	CCV05	CCV	03/21/24 16:29	CCV Fail For S	Bin	OK
60	CCB05	CCB05	CCB	03/21/24 16:33		Bin	OK

Instrument ID: P4

Daily Analysis Runlog For Sequence/QC Batch ID # LB129987

Review By		Review On					
Supervise By		Supervise On					
STD. NAME		STD REF.#					
ICAL Standard ICV Standard CCV Standard ICSA Standard CRI Standard LCS Standard Chk Standard							
61	LR1	LR1	HIGH STD	03/21/24 16:59		Bin	OK
62	LR2	LR2	HIGH STD	03/21/24 17:06		Bin	OK
63	P1816-01MS	PL-01-032024MS	MS	03/21/24 17:11		Bin	OK
64	P1816-01MSD	PL-01-032024MSD	MSD	03/21/24 17:16		Bin	OK
65	P1816-01A	PL-01-032024A	PS	03/21/24 17:20		Bin	OK
66	P1803-02	TRENCH-PIT	SAM	03/21/24 17:24		Bin	OK
67	P1807-01DL	3461DL	SAM	03/21/24 17:29		Bin	OK
68	P1807-02	3443	SAM	03/21/24 17:33		Bin	OK
69	P1807-03	3436-37	SAM	03/21/24 17:38		Bin	OK
70	P1807-04	3450-59	SAM	03/21/24 17:42		Bin	OK
71	CCV06	CCV06	CCV	03/21/24 17:52		Bin	OK
72	CCB06	CCB06	CCB	03/21/24 17:57		Bin	OK
73	P1807-05DL	3462DL	SAM	03/21/24 18:11		Bin	OK
74	P1807-05DUPDL	3462DUPDL	DUP	03/21/24 18:16		Bin	OK
75	P1807-05LDL	3462LDL	SD	03/21/24 18:20		Bin	OK
76	P1807-05MSDL	3462MSDL	MS	03/21/24 18:25		Bin	OK
77	P1807-05MSDDL	3462MSDDL	MSD	03/21/24 18:29		Bin	OK
78	P1807-05ADL	3462ADL	PS	03/21/24 18:43		Bin	OK
79	P1816-01DUP	PL-01-032024DUP	DUP	03/21/24 18:48		Bin	OK
80	P1816-01L	PL-01-032024L	SD	03/21/24 18:52		Bin	OK
81	CCV07	CCV07	CCV	03/21/24 18:56	CCV Fail For S	Bin	OK

Instrument ID: P4

Daily Analysis Runlog For Sequence/QC Batch ID # LB129987

Review By		Review On					
Supervise By		Supervise On					
STD. NAME	STD REF.#						
ICAL Standard ICV Standard CCV Standard ICSA Standard CRI Standard LCS Standard Chk Standard							

82	CCB07	CCB07	CCB	03/21/24 19:01		Bin	OK
83	P1814-01RE	TP-6RE	SAM	03/21/24 19:20	Not Required	Bin	Not Ok
84	P1815-01	OK-01-032024	SAM	03/21/24 19:24		Bin	OK
85	P1817-01	B-148-SB01	SAM	03/21/24 19:29		Bin	OK
86	P1817-02	B-148-SB03	SAM	03/21/24 19:33		Bin	OK
87	P1817-03	B-148-SB02	SAM	03/21/24 19:37		Bin	OK
88	P1821-01	72-12002	SAM	03/21/24 19:41		Bin	OK
89	P1786-02	TP-22	SAM	03/21/24 19:46		Bin	OK
90	P1616-01	196-BAY-AVENUE	SAM	03/21/24 19:50	CCV Fail For S	Bin	Not Ok
91	P1616-03	146-BAY-AVENUE	SAM	03/21/24 19:54	CCV Fail For S	Bin	Not Ok
92	P1792-01RE	72-11977RE	SAM	03/21/24 19:58	Not Required	Bin	Not Ok
93	CCV08	CCV08	CCV	03/21/24 20:03		Bin	OK
94	CCB08	CCB08	CCB	03/21/24 20:07		Bin	OK
95	P1793-02RE	OILY-DEBRIS-DRUM	MS	03/21/24 20:12	Not Required	Bin	Not Ok
96	P1810-01	342	SAM	03/21/24 20:16		Bin	OK
97	PB159695BL	PB159695BL	MB	03/21/24 20:21		Bin	OK
98	PB159695BS	PB159695BS	LCS	03/21/24 20:25		Bin	OK
99	PB159714BL	PB159714BL	MB	03/21/24 20:29		Bin	OK
100	PB159714BS	PB159714BS	LCS	03/21/24 20:34		Bin	OK
101	PB159676TB	PB159676TB	MB	03/21/24 20:38		Bin	OK
102	P1798-01	S-1	SAM	03/21/24 20:43		Bin	OK

Instrument ID: P4

Daily Analysis Runlog For Sequence/QC Batch ID # LB129987

Review By		Review On					
Supervise By		Supervise On					
STD. NAME	STD REF.#						
ICAL Standard ICV Standard CCV Standard ICSA Standard CRI Standard LCS Standard Chk Standard							

103	P1798-02	S-2	SAM	03/21/24 20:47		Bin	OK
104	P1798-03	S-3	SAM	03/21/24 20:52		Bin	OK
105	CCV09	CCV09	CCV	03/21/24 21:03		Bin	OK
106	CCB09	CCB09	CCB	03/21/24 21:08		Bin	OK
107	P1798-04	S-4	SAM	03/21/24 21:12		Bin	OK
108	P1798-05	S-5	SAM	03/21/24 21:17		Bin	OK
109	P1798-06RE	S-6RE	SAM	03/21/24 21:21	Not Required	Bin	Not Ok
110	P1798-07	S-7	SAM	03/21/24 21:26		Bin	OK
111	P1798-08	S-8	SAM	03/21/24 21:30		Bin	OK
112	P1798-09	S-9	SAM	03/21/24 21:35		Bin	OK
113	P1798-10	S-10	SAM	03/21/24 21:39		Bin	OK
114	P1798-11	S-11	SAM	03/21/24 21:44		Bin	OK
115	P1798-12	S-12	SAM	03/21/24 21:48		Bin	OK
116	P1798-13	S-13	SAM	03/21/24 21:53		Bin	OK
117	CCV10	CCV10	CCV	03/21/24 21:58		Bin	OK
118	CCB10	CCB10	CCB	03/21/24 22:02		Bin	OK
119	P1798-14	S-14	SAM	03/21/24 22:07		Bin	OK
120	P1798-15	S-15	SAM	03/21/24 22:11		Bin	OK
121	PB159715BL	PB159715BL	MB	03/21/24 22:16		Bin	OK
122	PB159715BS	PB159715BS	LCS	03/21/24 22:20		Bin	OK
123	PB159640BL	PB159640BL	MB	03/21/24 22:24		Bin	OK

Instrument ID: P4

Daily Analysis Runlog For Sequence/QC Batch ID # LB129987

Review By		Review On					
Supervise By		Supervise On					
STD. NAME		STD REF.#					
ICAL Standard ICV Standard CCV Standard ICSA Standard CRI Standard LCS Standard Chk Standard							

124	PB159640BS	PB159640BS	LCS	03/21/24 22:29		Bin	OK
125	P1784-01	MW-5B	SAM	03/21/24 22:33		Bin	OK
126	P1784-01DUP	MW-5BDUP	DUP	03/21/24 22:37		Bin	OK
127	P1784-01L	MW-5BL	SD	03/21/24 22:42		Bin	OK
128	P1784-01MS	MW-5BMS	MS	03/21/24 22:46		Bin	OK
129	CCV11	CCV11	CCV	03/21/24 22:51		Bin	OK
130	CCB11	CCB11	CCB	03/21/24 22:55		Bin	OK
131	P1784-01MSD	MW-5BMSD	MSD	03/21/24 23:00		Bin	OK
132	P1784-01A	MW-5BA	PS	03/21/24 23:04		Bin	OK
133	P1784-02	MW-7	SAM	03/21/24 23:08		Bin	OK
134	P1784-03	MW-3	SAM	03/21/24 23:13		Bin	OK
135	P1814-01	TP-6	SAM	03/21/24 23:17		Bin	OK
136	P1786-01MS	TP-15MS	MS	03/21/24 23:22	ICV Fail For S	Bin	Not Ok
137	P1786-01MSD	TP-15MSD	MSD	03/21/24 23:26	ICV Fail For S	Bin	Not Ok
138	P1786-01A	TP-15A	PS	03/21/24 23:30	ICV Fail For S	Bin	Not Ok
139	P1798-06	S-6	SAM	03/21/24 23:34		Bin	OK
140	PB159689TB	PB159689TB	MB	03/21/24 23:39		Bin	OK
141	CCV12	CCV12	CCV	03/21/24 23:43		Bin	OK
142	CCB12	CCB12	CCB	03/21/24 23:51		Bin	OK
143	P1827-01	72-11933	SAM	03/21/24 23:55		Bin	OK
144	P1827-01DL	72-11933DL	SAM	03/22/24	Not Required	Bin	Not Ok



Instrument ID: P4

Daily Analysis Runlog For Sequence/QC Batch ID # LB129987

Review By		Review On					
Supervise By		Supervise On					
STD. NAME		STD REF.#					
ICAL Standard ICV Standard CCV Standard ICSA Standard CRI Standard LCS Standard Chk Standard							
145	PB159637BL	PB159637BL	MB	03/22/24 00:10	ICV Fail For Fe	Bin	Not Ok
146	PB159637BS	PB159637BS	LCS	03/22/24 00:14	ICV Fail For Fe	Bin	Not Ok
147	P1788-01	PEZOMETER-200	SAM	03/22/24 00:18		Bin	OK
148	P1788-02	PEZOMETER-206	SAM	03/22/24 00:23		Bin	OK
149	CCV13	CCV13	CCV	03/22/24 00:37		Bin	OK
150	CCB13	CCB13	CCB	03/22/24 00:41		Bin	OK

Instrument ID: P4

Daily Analysis Runlog For Sequence/QC Batch ID # LB130043

Review By		Review On					
Supervise By		Supervise On					
STD. NAME		STD REF.#					
ICAL Standard ICV Standard CCV Standard ICSA Standard CRI Standard LCS Standard Chk Standard							

Sr#	SampleID	ClientID	QcType	Date	Comment	Operator	Status
1	S0	S0	CAL1	03/26/24 10:29		Bin	OK
2	S1	S1	CAL2	03/26/24 10:33		Bin	OK
3	S2	S2	CAL3	03/26/24 10:38		Bin	OK
4	S3	S3	CAL4	03/26/24 10:42		Bin	OK
5	S4	S4	CAL5	03/26/24 10:46		Bin	OK
6	S5	S5	CAL6	03/26/24 10:51		Bin	OK
7	ICV01	ICV01	ICV	03/26/24 11:09	ICV Fail for Mo,Ni,Na,Tl (95-105) (200.7) & Mo (90-110) (6010D)	Bin	OK
8	LLICV01	LLICV01	LLICV	03/26/24 11:14		Bin	OK
9	ICB01	ICB01	ICB	03/26/24 11:18		Bin	OK
10	CRI01	CRI01	CRDL	03/26/24 11:22		Bin	OK
11	ICSA01	ICSA01	ICSA	03/26/24 12:56		Bin	OK
12	ICSAB01	ICSAB01	ICSAB	03/26/24 13:03		Bin	OK
13	ICSADL	ICSADL	ICSA	03/26/24 13:07		Bin	OK
14	ICSABDL	ICSABDL	ICSAB	03/26/24 13:12		Bin	OK
15	CCV01	CCV01	CCV	03/26/24 13:26	CCV Fail For Mg,K	Bin	OK
16	CCB01	CCB01	CCB	03/26/24 13:30		Bin	OK
17	P1797-01	001-WILLETS-PT-BLV	SAM	03/26/24 13:43		Bin	OK
18	P1797-02	002-35TH-AVE(MAR)	SAM	03/26/24 13:48		Bin	OK
19	P1845-01	BELL-CON	SAM	03/26/24 13:52		Bin	OK

Instrument ID: P4

Daily Analysis Runlog For Sequence/QC Batch ID # LB130043

Review By		Review On					
Supervise By		Supervise On					
STD. NAME	STD REF.#						
ICAL Standard ICV Standard CCV Standard ICSA Standard CRI Standard LCS Standard Chk Standard							

20	P1832-01DL	600581536DL	SAM	03/26/24 13:56	10X For Fe	Bin	Confirms
21	P1844-01	BURL-OIL	SAM	03/26/24 14:01		Bin	OK
22	P1844-03	BURL-OILY-DEBRIS	SAM	03/26/24 14:05		Bin	OK
23	P1843-01	FREEDOM-OIL	SAM	03/26/24 14:09		Bin	OK
24	P1820-01DL2	WASTE-WATERDL2	SAM	03/26/24 14:14	10X For Fe	Bin	Confirms
25	P1830-02DL	VNJ-206DL	SAM	03/26/24 14:19	10X For Fe	Bin	Confirms
26	P1830-04DL	VNJ-226DL	SAM	03/26/24 14:23	10X For Fe	Bin	Confirms
27	CCV02	CCV02	CCV	03/26/24 15:07		Bin	OK
28	CCB02	CCB02	CCB	03/26/24 15:26		Bin	OK
29	P1828-04	TRE-1290	SAM	03/26/24 15:31		Bin	OK
30	P1831-01DL	CL-01-032124DL	SAM	03/26/24 15:35	10X For Ca,Fe,Mg	Bin	Confirms
31	P1847-01	SU-02-032224	SAM	03/26/24 15:40		Bin	OK
32	P1829-01DL	TR-05-031224DL	SAM	03/26/24 15:44	10X For Ca,Fe,Mg	Bin	Confirms
33	P1842-01	EO-01-03222024-A	SAM	03/26/24 15:48		Bin	OK
34	P1842-03	EO-01-03222024-B	SAM	03/26/24 15:53		Bin	OK
35	LR-1	LR-1	HIGH STD	03/26/24 15:57		Bin	OK
36	LR-2	LR-2	HIGH STD	03/26/24 16:02		Bin	OK
37	PB159696BL	PB159696BL	MB	03/26/24 16:16		Bin	OK
38	PB159696BS	PB159696BS	LCS	03/26/24 16:20		Bin	OK
39	CCV03	CCV03	CCV	03/26/24 16:36		Bin	OK
40	CCB03	CCB03	CCB	03/26/24 16:40		Bin	OK

Instrument ID: P4

Daily Analysis Runlog For Sequence/QC Batch ID # LB130043

Review By		Review On					
Supervise By		Supervise On					
STD. NAME	STD REF.#						
ICAL Standard ICV Standard CCV Standard ICSA Standard CRI Standard LCS Standard Chk Standard							

41	PB159759BL	PB159759BL	MB	03/26/24 16:45		Bin	OK
42	PB159759BS	PB159759BS	LCS	03/26/24 16:49		Bin	OK
43	PB159637BL	PB159637BL	MB	03/26/24 16:53	CCV Fail For Cd	Bin	Not Ok
44	PB159637BS	PB159637BS	LCS	03/26/24 16:58	CCV Fail For Cd	Bin	Not Ok
45	P1847-01DUP	SU-02-032224DUP	DUP	03/26/24 17:02		Bin	OK
46	P1847-01L	SU-02-032224L	SD	03/26/24 17:06		Bin	OK
47	P1847-01MS	SU-02-032224MS	MS	03/26/24 17:11		Bin	OK
48	P1847-01MSD	SU-02-032224MSD	MSD	03/26/24 17:15		Bin	OK
49	P1847-01A	SU-02-032224A	PS	03/26/24 17:19		Bin	OK
50	CCV04	CCV04	CCV	03/26/24 17:23	CCV Fail For Cd,Tl	Bin	OK
51	CCB04	CCB04	CCB	03/26/24 17:28		Bin	OK
52	P1828-04DUP	TRE-1290DUP	DUP	03/26/24 17:32		Bin	OK
53	P1828-04L	TRE-1290L	SD	03/26/24 17:37		Bin	OK
54	P1828-04MS	TRE-1290MS	MS	03/26/24 17:41		Bin	OK
55	P1828-04MSD	TRE-1290MSD	MSD	03/26/24 17:45		Bin	OK
56	P1828-04A	TRE-1290A	PS	03/26/24 17:50		Bin	OK
57	P1829-01DUPDL	TR-05-031224DUPDL	DUP	03/26/24 17:54	10X For Ca,Fe,Mg	Bin	Confirms
58	P1829-01LDL	TR-05-031224LDL	SD	03/26/24 17:58	50X For Fe	Bin	Confirms
59	P1829-01MSDL	TR-05-031224MSDL	MS	03/26/24 18:03	10X For Ca,Fe,Mg	Bin	Confirms
60	P1829-01MSDDL	TR-05-031224MSDDL	MSD	03/26/24 18:07	10X For Ca,Fe,Mg	Bin	Confirms
61	CCV05	CCV05	CCV	03/26/24 18:12		Bin	OK



Instrument ID: P4

Daily Analysis Runlog For Sequence/QC Batch ID # LB130043

Review By		Review On					
Supervise By		Supervise On					
STD. NAME		STD REF.#					
ICAL Standard ICV Standard CCV Standard ICSA Standard CRI Standard LCS Standard Chk Standard							
62	CCB05	CCB05	CCB	03/26/24 18:16		Bin	OK
63	P1714-04DL	WATER-TREATMENT	SAM	03/26/24 18:20	Not Use	Bin	Not Ok
64	P1714-04DUPDL	WATER-TREATMENT	DUP	03/26/24 18:25	10X For Na	Bin	Confirms
65	P1714-04LDL	WATER-TREATMENT	SD	03/26/24 18:30	50X For Na	Bin	Confirms
66	P1714-04MSDL	WATER-TREATMENT	MS	03/26/24 18:34	10X For Na	Bin	Confirms
67	P1714-04MSDDL	WATER-TREATMENT	MSD	03/26/24 18:39	10X For Na	Bin	Confirms
68	P1714-04ADL	WATER-TREATMENT	PS	03/26/24 18:43	Not Use	Bin	Not Ok
69	CCV06	CCV06	CCV	03/26/24 18:48	CCV Fail For Mg	Bin	OK
70	CCB06	CCB06	CCB	03/26/24 18:52		Bin	OK