



284 Sheffield Street, Mountainside, NJ 07092 Phone: 908 789 8900 Fax: 908 789 8922

IDENTIFICATION SUMMARY
FOR MULTICOMPONENT ANALYTES

SAMPLE NO.

PB159587BS

Contract: LIRO01

Lab Code: CHEM Case No.: P1747 SAS No.: P1747 SDG NO.: P1747

Lab Sample ID: PB159587BS Date(s) Analyzed: 03/14/2024 03/14/2024

Instrument ID (1): ECD_O Instrument ID (2): ECD_O

GC Column: (1): ZB-MR1 ID: 0.32 (mm) GC Column: (2): ZB-MR2 ID: 0.32 (mm)

Data file PO102513.D

ANALYTE	COL	RT	RT WINDOW		CONCENTRATION	MEAN CONCENTRATION	%RPD
			FROM	TO			
Aroclor-1016	1	5.412	5.362	5.462	5.31	5.30	1.9
COLUMN 1	2	5.434	5.384	5.484	5.35		
	3	5.495	5.445	5.545	5.37		
	4	5.592	5.542	5.642	5.31		
	5	5.882	5.832	5.932	5.02		
	1	4.434	4.384	4.484	5.33		
COLUMN 2	2	4.452	4.402	4.502	5.48	5.20	
	3	4.624	4.574	4.674	5.28		
	4	4.665	4.615	4.715	5.12		
	5	4.873	4.823	4.923	5.04		
Aroclor-1260	1	6.992	6.942	7.042	4.99	5.00	3.92
COLUMN 1	2	7.247	7.197	7.297	5.51		
	3	7.603	7.553	7.653	4.62		
	4	7.825	7.775	7.875	4.90		
	5	8.131	8.081	8.181	5.12		
	1	5.885	5.835	5.935	5.13		
COLUMN 2	2	6.071	6.021	6.121	5.60	5.20	
	3	6.22	6.17	6.27	5.17		
	4	6.685	6.635	6.735	4.83		
	5	6.926	6.876	6.976	5.45		

IDENTIFICATION SUMMARY
FOR MULTICOMPONENT ANALYTES

SAMPLE NO.

PB159587BSD

Contract: LIRO01

Lab Code: CHEM Case No.: P1747 SAS No.: P1747 SDG NO.: P1747

Lab Sample ID: PB159587BSD Date(s) Analyzed: 03/14/2024 03/14/2024

Instrument ID (1): ECD_O Instrument ID (2): ECD_O

GC Column: (1): ZB-MR1 ID: 0.32 (mm) GC Column: (2): ZB-MR2 ID: 0.32 (mm)

Data file PO102514.D

ANALYTE	COL	RT	RT WINDOW		CONCENTRATION	MEAN CONCENTRATION	%RPD
			FROM	TO			
Aroclor-1016	1	5.412	5.362	5.462	5.17	5.10	1.94
COLUMN 1	2	5.434	5.384	5.484	5.20		
	3	5.495	5.445	5.545	5.18		
	4	5.592	5.542	5.642	5.16		
	5	5.882	5.832	5.932	4.94		
COLUMN 2	1	4.434	4.384	4.484	5.24	5.20	
	2	4.452	4.402	4.502	5.39		
	3	4.623	4.573	4.673	5.23		
	4	4.665	4.615	4.715	5.08		
	5	4.872	4.822	4.922	4.99		
Aroclor-1260	1	6.993	6.943	7.043	4.96	5.00	3.92
COLUMN 1	2	7.247	7.197	7.297	5.45		
	3	7.602	7.552	7.652	4.62		
	4	7.826	7.776	7.876	4.86		
	5	8.131	8.081	8.181	5.04		
COLUMN 2	1	5.884	5.834	5.934	5.08	5.20	
	2	6.071	6.021	6.121	5.55		
	3	6.221	6.171	6.271	5.12		
	4	6.685	6.635	6.735	4.78		
	5	6.926	6.876	6.976	5.38		

IDENTIFICATION SUMMARY
FOR MULTICOMPONENT ANALYTES

SAMPLE NO.

PB159600BS

Contract: LIRO01

Lab Code: CHEM Case No.: P1747 SAS No.: P1747 SDG NO.: P1747

Lab Sample ID: PB159600BS Date(s) Analyzed: 03/15/2024 03/15/2024

Instrument ID (1): ECD_O Instrument ID (2): ECD_O

GC Column: (1): ZB-MR1 ID: 0.32 (mm) GC Column: (2): ZB-MR2 ID: 0.32 (mm)

Data file PO102527.D

ANALYTE	COL	RT	RT WINDOW		CONCENTRATION	MEAN CONCENTRATION	%RPD
			FROM	TO			
Aroclor-1016	1	5.411	5.361	5.461	0.059	0.053	12.22
COLUMN 1	2	5.434	5.384	5.484	0.055		
	3	5.494	5.444	5.544	0.052		
	4	5.591	5.541	5.641	0.052		
	5	5.882	5.832	5.932	0.047		
	1	4.433	4.383	4.483	0.066		
COLUMN 2	2	4.451	4.401	4.501	0.061	0.060	
	3	4.623	4.573	4.673	0.060		
	4	4.665	4.615	4.715	0.057		
	5	4.873	4.823	4.923	0.054		
Aroclor-1260	1	6.992	6.942	7.042	0.052	0.050	19.67
COLUMN 1	2	7.247	7.197	7.297	0.055		
	3	7.602	7.552	7.652	0.042		
	4	7.826	7.776	7.876	0.052		
	5	8.13	8.08	8.18	0.047		
	1	5.884	5.834	5.934	0.058		
COLUMN 2	2	6.071	6.021	6.121	0.067	0.060	
	3	6.221	6.171	6.271	0.068		
	4	6.685	6.635	6.735	0.052		
	5	6.926	6.876	6.976	0.057		

IDENTIFICATION SUMMARY
FOR MULTICOMPONENT ANALYTES

SAMPLE NO.

PB159600BSD

Contract: LIRO01

Lab Code: CHEM Case No.: P1747 SAS No.: P1747 SDG NO.: P1747

Lab Sample ID: PB159600BSD Date(s) Analyzed: 03/15/2024 03/15/2024

Instrument ID (1): ECD_O Instrument ID (2): ECD_O

GC Column: (1): ZB-MR1 ID: 0.32 (mm) GC Column: (2): ZB-MR2 ID: 0.32 (mm)

Data file PO102528.D

ANALYTE	COL	RT	RT WINDOW		CONCENTRATION	MEAN CONCENTRATION	%RPD
			FROM	TO			
Aroclor-1016	1	5.41	5.36	5.46	0.057	0.055	6.22
COLUMN 1	2	5.433	5.383	5.483	0.058		
	3	5.494	5.444	5.544	0.056		
	4	5.591	5.541	5.641	0.052		
	5	5.882	5.832	5.932	0.050		
	1	4.433	4.383	4.483	0.062		
COLUMN 2	2	4.451	4.401	4.501	0.059	0.058	
	3	4.623	4.573	4.673	0.058		
	4	4.664	4.614	4.714	0.057		
	5	4.872	4.822	4.922	0.054		
Aroclor-1260	1	6.993	6.943	7.043	0.055	0.051	16.16
COLUMN 1	2	7.246	7.196	7.296	0.055		
	3	7.602	7.552	7.652	0.044		
	4	7.826	7.776	7.876	0.053		
	5	8.13	8.08	8.18	0.050		
	1	5.884	5.834	5.934	0.057		
COLUMN 2	2	6.071	6.021	6.121	0.064	0.060	
	3	6.22	6.17	6.27	0.073		
	4	6.684	6.634	6.734	0.050		
	5	6.926	6.876	6.976	0.056		