



284 Sheffield Street, Mountainside, NJ 07092 Phone: 908 789 8900 Fax: 908 789 8922

CALIBRATION VERIFICATION SUMMARY

Contract: LIRO01

Lab Code: CHEM Case No.: P1747 SAS No.: P1747 SDG NO.: P1747

Continuing Calib Date: 03/14/2024 Initial Calibration Date(s): 03/12/2024 03/12/2024

Continuing Calib Time: 21:26 Initial Calibration Time(s): 15:06 22:50

GC Column: ZB-MR1 ID: 0.32 (mm)

COMPOUND		CCAL RT	AVG RT	RT WINDOW		DIFF RT
				FROM	TO	
Aroclor-1016-1	(1)	5.41	5.42	5.32	5.52	0.01
Aroclor-1016-2	(2)	5.43	5.44	5.34	5.54	0.01
Aroclor-1016-3	(3)	5.49	5.50	5.40	5.60	0.01
Aroclor-1016-4	(4)	5.59	5.60	5.50	5.70	0.01
Aroclor-1016-5	(5)	5.88	5.89	5.79	5.99	0.01
Aroclor-1260-1	(1)	6.99	7.00	6.90	7.10	0.01
Aroclor-1260-2	(2)	7.25	7.25	7.15	7.35	0.00
Aroclor-1260-3	(3)	7.60	7.61	7.51	7.71	0.01
Aroclor-1260-4	(4)	7.83	7.83	7.73	7.93	0.01
Aroclor-1260-5	(5)	8.13	8.14	8.04	8.24	0.01
Tetrachloro-m-xylene		4.27	4.27	4.17	4.37	0.00
Decachlorobiphenyl		9.81	9.81	9.71	9.91	0.00



284 Sheffield Street, Mountainside, NJ 07092 Phone: 908 789 8900 Fax: 908 789 8922

CALIBRATION VERIFICATION SUMMARY

Contract: LIRO01

Lab Code: CHEM Case No.: P1747 SAS No.: P1747 SDG NO.: P1747

Continuing Calib Date: 03/14/2024 Initial Calibration Date(s): 03/12/2024 03/12/2024

Continuing Calib Time: 21:26 Initial Calibration Time(s): 15:06 22:50

GC Column: ZB-MR2 ID: 0.32 (mm)

COMPOUND		CCAL RT	AVG RT	RT WINDOW FROM TO		DIFF RT
Aroclor-1016-1	(1)	4.43	4.44	4.34	4.54	0.01
Aroclor-1016-2	(2)	4.45	4.45	4.35	4.55	0.00
Aroclor-1016-3	(3)	4.62	4.63	4.53	4.73	0.01
Aroclor-1016-4	(4)	4.66	4.67	4.57	4.77	0.01
Aroclor-1016-5	(5)	4.87	4.88	4.78	4.98	0.01
Aroclor-1260-1	(1)	5.88	5.89	5.79	5.99	0.01
Aroclor-1260-2	(2)	6.07	6.08	5.98	6.18	0.01
Aroclor-1260-3	(3)	6.22	6.23	6.13	6.33	0.01
Aroclor-1260-4	(4)	6.69	6.69	6.59	6.79	0.01
Aroclor-1260-5	(5)	6.93	6.93	6.83	7.03	0.00
Tetrachloro-m-xylene		3.38	3.39	3.29	3.49	0.01
Decachlorobiphenyl		8.25	8.25	8.15	8.35	0.00