



284 Sheffield Street, Mountainside, NJ 07092 Phone: 908 789 8900 Fax: 908 789 8922

CALIBRATION VERIFICATION SUMMARY

Contract: LIRO01

Lab Code: CHEM Case No.: P1747 SAS No.: P1747 SDG NO.: P1747

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 03/12/2024 03/12/2024

Client Sample No.: CCAL01 Date Analyzed: 03/14/2024

Lab Sample No.: AR1660CCC500 Data File : PO102499.D Time Analyzed: 15:28

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	5.413	5.315	5.515	479.370	500.000	-4.1
Aroclor-1016-2	5.435	5.338	5.538	479.540	500.000	-4.1
Aroclor-1016-3	5.496	5.399	5.599	484.480	500.000	-3.1
Aroclor-1016-4	5.593	5.496	5.696	487.030	500.000	-2.6
Aroclor-1016-5	5.883	5.786	5.986	485.050	500.000	-3.0
Aroclor-1260-1	6.994	6.897	7.097	477.170	500.000	-4.6
Aroclor-1260-2	7.248	7.151	7.351	505.010	500.000	1.0
Aroclor-1260-3	7.603	7.506	7.706	503.610	500.000	0.7
Aroclor-1260-4	7.827	7.730	7.930	494.910	500.000	-1.0
Aroclor-1260-5	8.132	8.035	8.235	506.240	500.000	1.2
Decachlorobiphenyl	9.808	9.714	9.914	53.420	50.000	6.8
Tetrachloro-m-xylene	4.269	4.171	4.371	42.760	50.000	-14.5



284 Sheffield Street, Mountainside, NJ 07092 Phone: 908 789 8900 Fax: 908 789 8922

CALIBRATION VERIFICATION SUMMARY

Contract: LIRO01

Lab Code: CHEM Case No.: P1747 SAS No.: P1747 SDG NO.: P1747

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 03/12/2024 03/12/2024

Client Sample No.: CCAL01 Date Analyzed: 03/14/2024

Lab Sample No.: AR1660CCC500 Data File : PO102499.D Time Analyzed: 15:28

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.435	4.337	4.537	490.100	500.000	-2.0
Aroclor-1016-2	4.453	4.354	4.554	498.980	500.000	-0.2
Aroclor-1016-3	4.624	4.526	4.726	486.500	500.000	-2.7
Aroclor-1016-4	4.665	4.567	4.767	473.850	500.000	-5.2
Aroclor-1016-5	4.874	4.775	4.975	492.130	500.000	-1.6
Aroclor-1260-1	5.886	5.788	5.988	492.380	500.000	-1.5
Aroclor-1260-2	6.072	5.975	6.175	523.830	500.000	4.8
Aroclor-1260-3	6.222	6.125	6.325	489.760	500.000	-2.0
Aroclor-1260-4	6.686	6.589	6.789	513.100	500.000	2.6
Aroclor-1260-5	6.928	6.830	7.030	536.280	500.000	7.3
Decachlorobiphenyl	8.249	8.154	8.354	55.010	50.000	10.0
Tetrachloro-m-xylene	3.384	3.285	3.485	42.910	50.000	-14.2