

Surrogate Summary

SDG No.: **P1747**

Client: **LiRo Engineers, Inc.**

Analytical Method: **8082A**

Lab Sample ID	Client ID	Parameter	Column	Spike	Result	Rec	Qual	Limits	
								Low	High
I.BLK-PO102408.D	PIBLK-PO102408.D	Tetrachloro-m-xylene	1	20	17.9	90		60	140
		Decachlorobiphenyl	1	20	18.4	92		60	140
		Tetrachloro-m-xylene	2	20	17.6	88		60	140
		Decachlorobiphenyl	2	20	18.3	92		60	140
I.BLK-PO102500.D	PIBLK-PO102500.D	Tetrachloro-m-xylene	1	20	20.0	100		60	140
		Decachlorobiphenyl	1	20	26.1	131		60	140
		Tetrachloro-m-xylene	2	20	19.9	100		60	140
		Decachlorobiphenyl	2	20	27.5	137		60	140
PB159587BL	PB159587BL	Tetrachloro-m-xylene	1	20	19.1	96		21	155
		Decachlorobiphenyl	1	20	23.7	118		10	173
		Tetrachloro-m-xylene	2	20	18.8	94		21	155
		Decachlorobiphenyl	2	20	25.0	125		10	173
PB159587BS	PB159587BS	Tetrachloro-m-xylene	1	20	20.0	100		21	155
		Decachlorobiphenyl	1	20	24.2	121		10	173
		Tetrachloro-m-xylene	2	20	17.7	89		21	155
		Decachlorobiphenyl	2	20	25.7	128		10	173
PB159587BSD	PB159587BSD	Tetrachloro-m-xylene	1	20	19.5	97		21	155
		Decachlorobiphenyl	1	20	23.7	119		10	173
		Tetrachloro-m-xylene	2	20	17.5	88		21	155
		Decachlorobiphenyl	2	20	25.1	126		10	173
P1747-01	MW-01	Tetrachloro-m-xylene	1	20	19.2	96		21	155
		Decachlorobiphenyl	1	20	18.9	94		10	173
		Tetrachloro-m-xylene	2	20	19.1	96		21	155
		Decachlorobiphenyl	2	20	19.8	99		10	173
P1747-02	MW-01-DUP	Tetrachloro-m-xylene	1	20	19.0	95		21	155
		Decachlorobiphenyl	1	20	18.2	91		10	173
		Tetrachloro-m-xylene	2	20	19.6	98		21	155
		Decachlorobiphenyl	2	20	19.5	98		10	173
P1747-04	MW-02	Tetrachloro-m-xylene	1	20	19.7	98		21	155
		Decachlorobiphenyl	1	20	19.0	95		10	173
		Tetrachloro-m-xylene	2	20	19.5	97		21	155
		Decachlorobiphenyl	2	20	20.5	103		10	173
P1747-05	TWP-04	Tetrachloro-m-xylene	1	20	20.4	102		21	155
		Decachlorobiphenyl	1	20	21.0	105		10	173
		Tetrachloro-m-xylene	2	20	20.7	104		21	155
		Decachlorobiphenyl	2	20	22.7	113		10	173
I.BLK-PO102520.D	PIBLK-PO102520.D	Tetrachloro-m-xylene	1	20	20.7	104		60	140
		Decachlorobiphenyl	1	20	27.7	138		60	140
		Tetrachloro-m-xylene	2	20	20.9	104		60	140
		Decachlorobiphenyl	2	20	29.7	149	*	60	140
PB159600BS	PB159600BS	Tetrachloro-m-xylene	1	20	15.7	79		60	140
		Decachlorobiphenyl	1	20	20.7	103		60	140
		Tetrachloro-m-xylene	2	20	15.7	79		60	140



Surrogate Summary

SDG No.: P1747

Client: LiRo Engineers, Inc.

Analytical Method: 608.3 PCB

Lab Sample ID	Client ID	Parameter	Column	Spike	Result	Rec	Qual	Limits	
								Low	High
PB159600BS	PB159600BS	Decachlorobiphenyl	2	20	22.1	111		60	140
PB159600BSD	PB159600BSD	Tetrachloro-m-xylene	1	20	16.3	81		60	140
		Decachlorobiphenyl	1	20	20.5	103		60	140
		Tetrachloro-m-xylene	2	20	15.4	77		60	140
		Decachlorobiphenyl	2	20	22.3	111		60	140
PB159600BL	PB159600BL	Tetrachloro-m-xylene	1	20	15.7	79		60	140
		Decachlorobiphenyl	1	20	20.6	103		60	140
		Tetrachloro-m-xylene	2	20	15.6	78		60	140
		Decachlorobiphenyl	2	20	22.1	111		60	140
P1747-03	MW-01	Tetrachloro-m-xylene	1	20	10.8	54	*	60	140
		Decachlorobiphenyl	1	20	22.1	111		60	140
		Tetrachloro-m-xylene	2	20	12.8	64		60	140
		Decachlorobiphenyl	2	20	24.3	121		60	140
I.BLK-PO102532.D	PIBLK-PO102532.D	Tetrachloro-m-xylene	1	20	21.2	106		60	140
		Decachlorobiphenyl	1	20	28.0	140		60	140
		Tetrachloro-m-xylene	2	20	21.2	106		60	140
		Decachlorobiphenyl	2	20	29.7	149	*	60	140