



284 Sheffield Street, Mountainside, NJ 07092 Phone: 908 789 8900 Fax: 908 789 8922

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.: P1747

Client: LiRo Engineers, Inc.

Analytical Method: 8082A

Datafile : PO102513.D

Lab Sample ID	Parameter	Spike	Result	Units	Rec	RPD	Qual	RPD	Low	Limits	RPD
								Qual		High	
PB159587BS	AR1016	5	5.30	ug/L	106				61	112	
	AR1260	5	5.20	ug/L	104				66	113	