



284 Sheffield Street, Mountainside, NJ 07092 Phone: 908 789 8900 Fax: 908 789 8922

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.: P1747

Client: LiRo Engineers, Inc.

Analytical Method: 8082A

Datafile : PO102514.D

Lab Sample ID	Parameter	Spike	Result	Units	Rec	RPD	Qual	RPD	Low	Limits	RPD
								Qual		High	
PB159587BSD	AR1016	5	5.20	ug/L	104	2			61	112	20
	AR1260	5	5.20	ug/L	104	0			66	113	20