



284 Sheffield Street, Mountainside, NJ 07092 Phone: 908 789 8900 Fax: 908 789 8922

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.: P1747

Client: LiRo Engineers, Inc.

Analytical Method: 608.3 PCB

Datafile : PO102528.D

Lab Sample ID	Parameter	Spike	Result	Units	Rec	RPD	Qual	RPD	Low	Limits	RPD
								Qual		High	
PB159600BSD	AROCLOR 1016	0.05	0.055	ug/L	109	3			50	140	20
	AROCLOR 1260	0.05	0.051	ug/L	102	3			8	140	20