



284 Sheffield Street, Mountainside, NJ 07092 Phone: 908 789 8900 Fax: 908 789 8922

INITIAL CALIBRATION OF MULTICOMPONENT ANALYTES

Contract: LIRO01

Lab Code: CHEM Case No.: P1747 SAS No.: P1747 SDG NO.: P1747

Instrument ID: ECD_L Date(s) Analyzed: 03/13/2024 03/13/2024

GC Column: ZB-MR2 ID: 0.32 (mm)

COMPOUND	AMOUNT (ng)	PEAK	RT	RT WINDOW		CALIBRATION FACTOR
				FROM	TO	
Toxaphene	500	1	5.99	5.89	6.09	15819600
		2	6.58	6.48	6.68	45191500
		3	7.21	7.11	7.31	123592000
		4	7.30	7.20	7.40	90671800
		5	7.73	7.63	7.83	68270000



284 Sheffield Street, Mountainside, NJ 07092 Phone: 908 789 8900 Fax: 908 789 8922

INITIAL CALIBRATION OF MULTICOMPONENT ANALYTES

Contract: LIRO01

Lab Code: CHEM Case No.: P1747 SAS No.: P1747 SDG NO.: P1747

Instrument ID: ECD_L Date(s) Analyzed: 03/13/2024 03/13/2024

GC Column: ZB-MR1 ID: 0.32 (mm)

COMPOUND	AMOUNT (ng)	PEAK	RT	RT WINDOW		CALIBRATION FACTOR
				FROM	TO	
Toxaphene	500	1	5.46	5.36	5.56	13114100
		2	5.82	5.72	5.92	13971000
		3	6.75	6.65	6.85	45709900
		4	6.88	6.78	6.98	62869100
		5	7.20	7.10	7.30	26040400