



284 Sheffield Street, Mountainside, NJ 07092 Phone: 908 789 8900 Fax: 908 789 8922

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.: P1747

Client: LiRo Engineers, Inc.

Analytical Method: 8081B

Datafile : PL088617.D

Lab Sample ID	Parameter	Spike	Result	Units	Rec	RPD	Qual	RPD		Limits	
								Qual	Low	High	RPD
PB159588BSD	alpha-BHC	0.5	0.48	ug/L	97	5			85	130	20
	beta-BHC	0.5	0.47	ug/L	94	4			83	126	20
	delta-BHC	0.5	0.44	ug/L	89	4			69	141	20
	gamma-BHC (Lindane)	0.5	0.47	ug/L	95	4			82	129	20
	Heptachlor	0.5	0.49	ug/L	97	4			79	127	20
	Aldrin	0.5	0.47	ug/L	94	4			79	126	20
	Heptachlor epoxide	0.5	0.48	ug/L	96	4			81	124	20
	Endosulfan I	0.5	0.48	ug/L	97	4			85	122	20
	Dieldrin	0.5	0.49	ug/L	97	4			83	125	20
	4,4'-DDE	0.5	0.48	ug/L	97	4			80	127	20
	Endrin	0.5	0.48	ug/L	96	5			81	128	20
	Endosulfan II	0.5	0.48	ug/L	96	4			82	123	20
	4,4'-DDD	0.5	0.48	ug/L	96	4			77	131	20
	Endosulfan sulfate	0.5	0.48	ug/L	96	3			76	129	20
	4,4'-DDT	0.5	0.49	ug/L	98	4			80	133	20
	Methoxychlor	0.5	0.46	ug/L	93	3			76	137	20
	Endrin ketone	0.5	0.48	ug/L	96	4			80	131	20
	Endrin aldehyde	0.5	0.50	ug/L	100	3			82	127	20
	alpha-Chlordane	0.5	0.49	ug/L	99	4			82	125	20
	gamma-Chlordane	0.5	0.50	ug/L	101	4			82	125	20