

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF022924\
 Data File : BF137447.D
 Acq On : 29 Feb 2024 17:44
 Operator : CG\JU
 Sample : SSTDICV040
 Misc :
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 ICVBF022924

Quant Time: Mar 01 03:42:02 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF022924.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Mar 01 03:39:04 2024
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	20.000	20.000	0.0	96	0.08
2	1,4-Dioxane	40.000	39.925	0.2	93	0.15
3	Pyridine	40.000	42.971	-7.4	94	0.19
4	n-Nitrosodimethylamine	40.000	41.212	-3.0	98	0.19
5 S	2-Fluorophenol	80.000	83.353	-4.2	96	0.10
6	Aniline	40.000	41.863	-4.7	95	0.09
7 S	Phenol-d6	80.000	81.205	-1.5	96	0.08
8	2-Chlorophenol	40.000	41.043	-2.6	97	0.08
9	Benzaldehyde	40.000	43.144	-7.9	96	0.09
10 C	Phenol	40.000	41.669	-4.2	96	0.08
11	bis(2-Chloroethyl)ether	40.000	40.426	-1.1	92	0.08
12	1,3-Dichlorobenzene	40.000	39.963	0.1	95	0.08
13 C	1,4-Dichlorobenzene	40.000	40.018	-0.0	95	0.08
14	1,2-Dichlorobenzene	40.000	40.423	-1.1	97	0.08
15	Benzyl Alcohol	40.000	42.931	-7.3	96	0.08
16	2,2'-oxybis(1-Chloropropane	40.000	39.912	0.2	95	0.08
17	2-Methylphenol	40.000	40.227	-0.6	95	0.08
18	Hexachloroethane	40.000	41.467	-3.7	97	0.08
19 P	n-Nitroso-di-n-propylamine	40.000	39.231	1.9	95	0.08
20	3+4-Methylphenols	40.000	40.534	-1.3	96	0.08
21 I	Naphthalene-d8	20.000	20.000	0.0	97	0.08
22	Acetophenone	40.000	39.879	0.3	97	0.08
23 S	Nitrobenzene-d5	80.000	81.270	-1.6	96	0.08
24	Nitrobenzene	40.000	41.647	-4.1	99	0.08
25	Isophorone	40.000	39.625	0.9	96	0.08
26 C	2-Nitrophenol	40.000	43.344	-8.4	101	0.08
27	2,4-Dimethylphenol	40.000	40.463	-1.2	96	0.08
28	bis(2-Chloroethoxy)methane	40.000	40.695	-1.7	97	0.08
29 C	2,4-Dichlorophenol	40.000	41.381	-3.5	97	0.08
30	1,2,4-Trichlorobenzene	40.000	40.223	-0.6	97	0.08
31	Naphthalene	40.000	39.355	1.6	97	0.08
32	Benzoic acid	40.000	38.580	3.6	100	0.08
33	4-Chloroaniline	40.000	41.650	-4.1	96	0.08
34 C	Hexachlorobutadiene	40.000	39.843	0.4	96	0.08
35	Caprolactam	40.000	41.651	-4.1	98	0.08
36 C	4-Chloro-3-methylphenol	40.000	40.697	-1.7	97	0.08
37	2-Methylnaphthalene	40.000	39.935	0.2	97	0.08
38	1-Methylnaphthalene	40.000	39.428	1.4	97	0.08
39 I	Acenaphthene-d10	20.000	20.000	0.0	97	0.08
40	1,2,4,5-Tetrachlorobenzene	40.000	40.358	-0.9	97	0.08
41 P	Hexachlorocyclopentadiene	40.000	41.513	-3.8	108	0.08
42 S	2,4,6-Tribromophenol	80.000	81.217	-1.5	98	0.08
43 C	2,4,6-Trichlorophenol	40.000	40.691	-1.7	96	0.08
44	2,4,5-Trichlorophenol	40.000	41.118	-2.8	97	0.07
45 S	2-Fluorobiphenyl	80.000	76.340	4.6	95	0.07
46	1,1'-Biphenyl	40.000	40.167	-0.4	97	0.07
47	2-Chloronaphthalene	40.000	39.747	0.6	96	0.07

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF022924\
 Data File : BF137447.D
 Acq On : 29 Feb 2024 17:44
 Operator : CG\JU
 Sample : SSTDICV040
 Misc :
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 ICVBF022924

Quant Time: Mar 01 03:42:02 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF022924.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Mar 01 03:39:04 2024
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
48	2-Nitroaniline	40.000	43.934	-9.8	100	0.08
49	Acenaphthylene	40.000	39.534	1.2	95	0.08
50	Dimethylphthalate	40.000	39.579	1.1	97	0.08
51	2,6-Dinitrotoluene	40.000	41.886	-4.7	98	0.08
52 C	Acenaphthene	40.000	39.348	1.6	96	0.08
53	3-Nitroaniline	40.000	43.361	-8.4	98	0.08
54 P	2,4-Dinitrophenol	40.000	40.428	-1.1	101	0.08
55	Dibenzofuran	40.000	40.123	-0.3	98	0.08
56 P	4-Nitrophenol	40.000	41.613	-4.0	101	0.08
57	2,4-Dinitrotoluene	40.000	43.620	-9.0	100	0.08
58	Fluorene	40.000	38.909	2.7	97	0.08
59	2,3,4,6-Tetrachlorophenol	40.000	41.009	-2.5	98	0.08
60	Diethylphthalate	40.000	39.937	0.2	97	0.08
61	4-Chlorophenyl-phenylether	40.000	38.430	3.9	97	0.08
62	4-Nitroaniline	40.000	43.902	-9.8	103	0.08
63	Azobenzene	40.000	40.315	-0.8	97	0.08
64 I	Phenanthrene-d10	20.000	20.000	0.0	98	0.08
65	4,6-Dinitro-2-methylphenol	40.000	43.367	-8.4	101	0.08
66 c	n-Nitrosodiphenylamine	40.000	40.298	-0.7	100	0.08
67	4-Bromophenyl-phenylether	40.000	40.510	-1.3	97	0.08
68	Hexachlorobenzene	40.000	40.339	-0.8	100	0.08
69	Atrazine	40.000	40.634	-1.6	99	0.08
70 C	Pentachlorophenol	40.000	42.470	-6.2	99	0.08
71	Phenanthrene	40.000	39.294	1.8	99	0.08
72	Anthracene	40.000	39.514	1.2	99	0.08
73	Carbazole	40.000	40.364	-0.9	99	0.08
74	Di-n-butylphthalate	40.000	40.758	-1.9	99	0.08
75 C	Fluoranthene	40.000	39.697	0.8	100	0.08
76 I	Chrysene-d12	20.000	20.000	0.0	103	0.08
77	Benzydine	40.000	41.378	-3.4	92	0.08
78	Pyrene	40.000	40.735	-1.8	100	0.08
79 S	Terphenyl-d14	80.000	80.502	-0.6	101	0.08
80	Butylbenzylphthalate	40.000	44.194	-10.5	108	0.08
81	Benzo(a)anthracene	40.000	39.894	0.3	102	0.08
82	3,3'-Dichlorobenzidine	40.000	42.162	-5.4	97	0.08
83	Chrysene	40.000	40.308	-0.8	102	0.08
84	Bis(2-ethylhexyl)phthalate	40.000	44.058	-10.1	107	0.08
85 c	Di-n-octyl phthalate	40.000	38.294	4.3	98	0.09
86 I	Perylene-d12	20.000	20.000	0.0	98	0.11
87	Indeno(1,2,3-cd)pyrene	40.000	41.289	-3.2	96	0.16
88	Benzo(b)fluoranthene	40.000	39.969	0.1	97	0.11
89	Benzo(k)fluoranthene	40.000	41.030	-2.6	104	0.10
90 C	Benzo(a)pyrene	40.000	41.142	-2.9	97	0.12
91	Dibenzo(a,h)anthracene	40.000	41.702	-4.3	96	0.17
92	Benzo(g,h,i)perylene	40.000	41.441	-3.6	96	0.16

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF022924\
Data File : BF137447.D
Acq On : 29 Feb 2024 17:44
Operator : CG\JU
Sample : SSTDICV040
Misc :
ALS Vial : 10 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
ICVBF022924

Quant Time: Mar 01 03:42:02 2024
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF022924.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Mar 01 03:39:04 2024
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	Amount	Calc.	%Dev	Area%	Dev(min)
----------	--------	-------	------	-------	----------

(#) = Out of Range SPCC's out = 0 CCC's out = 0