



284 Sheffield Street, Mountainside, NJ 07092 Phone: 908 789 8900 Fax: 908 789 8922

Surrogate Summary

SW-846

SDG No.: P1747

Client: LiRo Engineers, Inc.

Analytical Method: 625.1

Lab Sample ID	Client ID	Parameter	Spike (PPM)	Result (PPM)	Recovery (%)	Qual	Limits (%)	
							Low	High
P1747-03	MW-01	2-Fluorophenol	100	41.9	42	*	60	140
		Phenol-d6	100	22.8	23	*	60	140
		Nitrobenzene-d5	100	87.3	87		60	140
		2-Fluorobiphenyl	100	86.6	87		60	140
		2,4,6-Tribromophenol	100	98.9	99		60	140
P1747-03RE	MW-01RE	Terphenyl-d14	100	62.4	62		60	140
		2-Fluorophenol	100	40.2	40	*	60	140
		Phenol-d6	100	22.4	22	*	60	140
		Nitrobenzene-d5	100	94.8	95		60	140
		2-Fluorobiphenyl	100	96.0	96		60	140
PB159613BL	PB159613BL	2,4,6-Tribromophenol	100	98.4	98		60	140
		Terphenyl-d14	100	84.2	84		60	140
		2-Fluorophenol	100	91.8	92		60	140
		Phenol-d6	100	84.3	84		60	140
		Nitrobenzene-d5	100	84.1	84		60	140
PB159613BS	PB159613BS	2-Fluorobiphenyl	100	85.8	86		60	140
		2,4,6-Tribromophenol	100	85.2	85		60	140
		Terphenyl-d14	100	77.7	78		60	140
		2-Fluorophenol	100	86.3	86		60	140
		Phenol-d6	100	80.5	81		60	140
PB159613BSD	PB159613BSD	Nitrobenzene-d5	100	82.2	82		60	140
		2-Fluorobiphenyl	100	82.1	82		60	140
		2,4,6-Tribromophenol	100	90.7	91		60	140
		Terphenyl-d14	100	76.6	77		60	140
		2-Fluorophenol	100	84.1	84		60	140
		Phenol-d6	100	78.3	78		60	140
		Nitrobenzene-d5	100	77.7	78		60	140
		2-Fluorobiphenyl	100	79.4	79		60	140
		2,4,6-Tribromophenol	100	87.2	87		60	140
		Terphenyl-d14	100	73.3	73		60	140