



284 Sheffield Street, Mountainside, NJ 07092 Phone: 908 789 8900 Fax: 908 789 8922

8B

SEMIVOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH

Lab Code: CHEM Case No.: P1747 SAS No.: P1747 SDG NO.: P1747

EPA Sample No.: SSTDCCC040 Date Analyzed: 03/15/2024

Lab File ID: BF137610.D Time Analyzed: 15:34

Instrument ID: BNA_F GC Column: DB-UI ID: 0.18 (mm)

	IS1 (DCB) AREA #	RT #	IS2 (NPT) AREA #	RT #	IS3 (ANT) AREA #	RT #
12 HOUR STD	156491	6.728	579710	8.01	317075	9.76
UPPER LIMIT	312982	7.228	1159420	8.51	634150	10.263
LOWER LIMIT	78245.5	6.228	289855	7.51	158538	9.263
EPA SAMPLE NO.						
01 PB159613BL	164876	6.73	640676	8.00	350327	9.76
02 PB159613BS	174702	6.73	662837	8.01	361357	9.76
03 MW-01	122193	6.72	480505	8.00	271859	9.76
04 PB159613BSD	177128	6.73	685431	8.01	367798	9.76

IS1 (DCB) = 1,4-Dichlorobenzene-d4

IS2 (NPT) = Naphthalene-d8

IS3 (ANT) = Acenaphthene-d10

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT UPPER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.